

LB129850


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

61187



A= 0.0000e+000

B= 1.6614e-004

C= 5.6465e-002

Rho= 0.9988390

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/1
0.0	0.000	0.055	0.055	-6	0.000	-6					
0.2	0.200	0.229	0.029	1039	0.0 %	1039					15
2.5	2.500	2.490	-0.010	14646	0.0 %	14646					0
5.0	5.000	5.067	0.067	30159	0.0 %	30159					1
7.5	7.500	7.137	-0.363	42618	0.0 %	42618					-5
10.0	10.000	10.222	0.222	61187	0.0 %	61187					2

PB129850 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Mar 2024 10:06:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-6	-	0.0000	7471B	PPB	12 Mar 2024 10:13:40	S	Std
0.2 - 1	50.2	1039	-	0.2000	7471B	PPB	12 Mar 2024 10:15:58	S	Std
2.5 - 1	52.5	14646	-	2.5000	7471B	PPB	12 Mar 2024 10:18:16	S	Std
5.0 - 1	55	30159	-	5.0000	7471B	PPB	12 Mar 2024 10:23:18	S	Std
7.5 - 1	57.5	42618	-	7.5000	7471B	PPB	12 Mar 2024 10:25:37	S	Std
10.0 - 1	510	61187	-	10.0000	7471B	PPB	12 Mar 2024 10:30:49	S	Std
ICV44 - 1	ICV44	23776	4.0065	-	7471B	PPB	12 Mar 2024 10:33:39	U	SMPL
ICB44 - 1	ICB44	-279	0.0101	-	7471B	PPB	12 Mar 2024 10:35:56	U	SMPL
CCV54 - 1	CCV54	29799	5.0072	-	7471B	PPB	12 Mar 2024 10:38:17	U	SMPL
CCB54 - 1	CCB54	-315	0.0041	-	7471B	PPB	12 Mar 2024 10:40:32	U	SMPL
CRA - 1	CRA	969	0.2175	-	7471B	PPB	12 Mar 2024 10:42:55	U	SMPL
HighStd - 1	HighStd	60622	10.1280	-	7471B	PPB	12 Mar 2024 10:45:10	U	SMPL
ChkStd - 1	ChkStd	39664	6.6461	-	7471B	PPB	12 Mar 2024 10:47:25	U	SMPL
PB159551BL - 1	PBS	-242	0.0163	-	7471B	PPB	12 Mar 2024 10:49:55	U	SMPL
PB159551BS - 1	LCSS	23120	3.8975	-	7471B	PPB	12 Mar 2024 10:54:55	U	SMPL
P1706-01 - 1	AUD-1333	1297	0.2719	-	7471B	PPB	12 Mar 2024 10:57:11	U	SMPL
P1710-01 - 1	TR-04-030724	10289	1.7658	-	7471B	PPB	12 Mar 2024 10:59:33	U	SMPL
P1720-01 - 1	RT4196	505	0.1404	-	7471B	PPB	12 Mar 2024 11:01:53	U	SMPL
P1720-03 - 1	RT3471	16126	2.7356	-	7471B	PPB	12 Mar 2024 11:04:13	U	SMPL
P1724-01 - 1	13925	531	0.1447	-	7471B	PPB	12 Mar 2024 11:06:28	U	SMPL
CCV55 - 1	CCV55	29044	4.8817	-	7471B	PPB	12 Mar 2024 11:17:04	U	SMPL
CCB55 - 1	CCB55	-261	0.0131	-	7471B	PPB	12 Mar 2024 11:19:19	U	SMPL
P1724-02 - 1	10252	7160	1.2460	-	7471B	PPB	12 Mar 2024 11:21:46	U	SMPL
P1724-03 - 1	90	451	0.1314	-	7471B	PPB	12 Mar 2024 11:24:01	U	SMPL
P1725-01 - 1	RT-3860	922	0.2096	-	7471B	PPB	12 Mar 2024 11:26:20	U	SMPL
P1725-01DUP - 1	RT-3860DUP	981	0.2194	-	7471B	PPB	12 Mar 2024 11:28:35	U	SMPL
P1725-01MS - 1	RT-3860MS	20318	3.4320	-	7471B	PPB	12 Mar 2024 11:30:51	U	SMPL
P1725-01MSD - 1	RT-3860MSD	19505	3.2970	-	7471B	PPB	12 Mar 2024 11:33:08	U	SMPL
P1725-01LX5 - 1		-76	0.0438	-	7471B	PPB	12 Mar 2024 11:35:30	U	SMPL
P1725-01A - 1		20151	3.4043	-	7471B	PPB	12 Mar 2024 11:37:53	U	SMPL
CCV56 - 1	CCV56	28000	4.7083	-	7471B	PPB	12 Mar 2024 11:40:10	U	SMPL
CCB56 - 1	CCB56	-263	0.0128	-	7471B	PPB	12 Mar 2024 11:42:31	U	SMPL