

LB129852


SFAM01.1

INSTRUMENT ID: CX1

Linear

 μ Abs.:

51625



A= 0.0000e+000

B= 1.9296e-004 *slope*C= 1.6001e-002 *y-intercept*

Rho= 0.9999029

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>%d)</i>
0.00	0.000	0.029	0.029	69	0.000	69					
0.05	0.050					442	0				<i>=</i>
0.20	0.200	0.173	-0.027	816	0.0 %	816					<i>-14</i>
2.50	2.500	2.530	0.030	13028	0.0 %	13028					<i>1</i>
5.00	5.000	4.914	-0.086	25386	0.0 %	25386					<i>-2</i>
7.50	7.500	7.575	0.075	39174	0.0 %	39174					<i>1</i>
10.0	10.000	9.978	-0.022	51629	0.0 %	51629					<i>0</i>

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	69	-		0 SFAM01.1	PPB	3/12/2024 15:24	Std
	0.2 S01	816	-		0.2 SFAM01.1	PPB	3/12/2024 15:27	Std
	2.5 S02	13028	-		2.5 SFAM01.1	PPB	3/12/2024 15:31	Std
	5 S03	25386	-		5 SFAM01.1	PPB	3/12/2024 15:35	Std
	7.5 S04	39174	-		7.5 SFAM01.1	PPB	3/12/2024 15:38	Std
	10 S05	51629	-		10 SFAM01.1	PPB	3/12/2024 15:46	Std
ICV045	ICV045	20282	3.9295 -		SFAM01.1	PPB	3/12/2024 15:48	SMPL
ICB045	ICB045	-146	-0.0122 -		SFAM01.1	PPB	3/12/2024 15:51	SMPL
CCV057	CCV057	25299	4.8976 -		SFAM01.1	PPB	3/12/2024 15:53	SMPL
CCB057	CCB057	-138	-0.0106 -		SFAM01.1	PPB	3/12/2024 15:55	SMPL
PB159557BL	PBW57	184	0.0515 -		SFAM01.1	PPB	3/12/2024 15:58	SMPL
P1609-03	MDCNH0	-177	-0.0182 -		SFAM01.1	PPB	3/12/2024 16:02	SMPL
P1609-05	MDCNH0S	5723	1.1203 -		SFAM01.1	PPB	3/12/2024 16:10	SMPL
P1609-06	MDCNH2	18	0.0195 -		SFAM01.1	PPB	3/12/2024 16:13	SMPL
P1609-07	MDCNH6	23	0.0204 -		SFAM01.1	PPB	3/12/2024 16:15	SMPL
P1609-08	MDCNH8	26	0.021 -		SFAM01.1	PPB	3/12/2024 16:17	SMPL
P1609-09	MDCNI0	39	0.0235 -		SFAM01.1	PPB	3/12/2024 16:19	SMPL
P1609-10	MDCNG0	121	0.0393 -		SFAM01.1	PPB	3/12/2024 16:22	SMPL
P1609-11	MDCNG2	222	0.0588 -		SFAM01.1	PPB	3/12/2024 16:24	SMPL
P1609-12	MDCNG4	99	0.0351 -		SFAM01.1	PPB	3/12/2024 16:26	SMPL
P1609-13	MDCNG6	64	0.0284 -		SFAM01.1	PPB	3/12/2024 16:28	SMPL
P1609-14	MDCNG8	151	0.0451 -		SFAM01.1	PPB	3/12/2024 16:31	SMPL
PB159558BL	PBW558	17	0.0193 -		SFAM01.1	PPB	3/12/2024 16:33	SMPL
P1643-01	MDCNG1	40	0.0237 -		SFAM01.1	PPB	3/12/2024 16:35	SMPL
P1643-02	MDCNG3	39	0.0235 -		SFAM01.1	PPB	3/12/2024 16:37	SMPL
P1643-03	MDCNG5	48	0.0253 -		SFAM01.1	PPB	3/12/2024 16:40	SMPL
P1643-04	MDCNG7	28	0.0214 -		SFAM01.1	PPB	3/12/2024 16:42	SMPL
P1643-05	MDCNG9	78	0.0311 -		SFAM01.1	PPB	3/12/2024 16:44	SMPL
CCV058	CCV058	24517	4.7467 -		SFAM01.1	PPB	3/12/2024 16:47	SMPL
CCB058	CCB058	-80	0.0006 -		SFAM01.1	PPB	3/12/2024 16:49	SMPL
P1643-06	MDCNH1	51	0.0258 -		SFAM01.1	PPB	3/12/2024 16:51	SMPL
P1643-07	MDCNH1D	43	0.0243 -		SFAM01.1	PPB	3/12/2024 16:53	SMPL
P1643-08	MDCNH1S	5280	1.0348 -		SFAM01.1	PPB	3/12/2024 16:56	SMPL
P1643-09	MDCNH3	-16	0.0129 -		SFAM01.1	PPB	3/12/2024 16:58	SMPL
P1643-10	MDCNH7	32	0.0222 -		SFAM01.1	PPB	3/12/2024 17:00	SMPL
P1643-11	MDCNH9	16	0.0191 -		SFAM01.1	PPB	3/12/2024 17:02	SMPL
P1643-12	MDCNI1	42	0.0241 -		SFAM01.1	PPB	3/12/2024 17:05	SMPL
P1609-04	MDCNH0D	63	0.0282 -		SFAM01.1	PPB	3/12/2024 17:07	SMPL
P1609-02	DCNN4	41757	8.0733 -		SFAM01.1	PPB	3/12/2024 17:12	SMPL
CCV059	CCV059	23356	4.5227 -		SFAM01.1	PPB	3/12/2024 17:14	SMPL
CCB059	CCB059	-139	-0.0108 -		SFAM01.1	PPB	3/12/2024 17:16	SMPL