



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB129855

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/12/2024	14:35
2	CAL2	1	Water	NA	NA	20.3	7.00	03/12/2024	14:36
3	CAL3	1	Water	NA	NA	20.3	10.02	03/12/2024	14:39
4	ICV	1	Water	NA	NA	20.5	7.01	03/12/2024	14:40
5	CCV1	1	Water	NA	NA	20.4	2.01	03/12/2024	14:45
6	P1725-02	1	Solid	20.02	20	22.6	6.34	03/12/2024	14:50
7	P1733-04	1	Solid	20.03	20	25.2	6.44	03/12/2024	15:00
8	P1733-08	1	Solid	20.02	20	24.2	6.35	03/12/2024	15:10
9	P1733-12	1	Solid	20.03	20	24.8	6.75	03/12/2024	15:15
10	P1733-12DUP	1	Solid	20.04	20	24.9	6.76	03/12/2024	15:16
11	CCV2	1	Water	NA	NA	20.3	12.02	03/12/2024	15:20