


LB129884

245.1

INSTRUMENT ID: 042

Linear



A= 0.0000e+000

B= 2.5552e-004

C= -6.2975e-003

Rho= 0.9999079

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | U, I) |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|-------|
| 0.0    | 0.000  | 0.006  | 0.006  | 47    | 0.000      | 47    |       |       |       |       | -     |
| 0.05   | 0.050  | 0.047  | -0.003 | 210   | 0.0 %      | 210   | 0     |       |       |       | -6    |
| 0.2    | 0.200  | 0.216  | 0.016  | 871   | 0.0 %      | 871   |       |       |       |       | 8     |
| 2.5    | 2.500  | 2.458  | -0.042 | 9644  | 0.0 %      | 9644  |       |       |       |       | -2    |
| 5.0    | 5.000  | 5.075  | 0.075  | 19885 | 0.0 %      | 19885 |       |       |       |       | 2     |
| 7.5    | 7.500  | 7.408  | -0.092 | 29016 | 0.0 %      | 29016 |       |       |       |       | 1     |
| 10.0   | 10.000 | 10.040 | 0.040  | 39318 | 0.0 %      | 39318 |       |       |       |       | 0     |

## LB129864 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 13 Mar 2024 11:34:21

| Sample ID       | Extended ID               | μ Abs | Conc    | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|---------------------------|-------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | 50                        | 47    | -       | 0.0000   | 245.1  | PPB   | 13 Mar 2024 11:53:47 | S    | Std  |
| 0.05 - 1        | 50.05                     | 210   | -       | 0.0500   | 245.1  | PPB   | 13 Mar 2024 11:59:18 | S    | Std  |
| 0.2 - 1         | 50.2                      | 871   | -       | 0.2000   | 245.1  | PPB   | 13 Mar 2024 12:01:35 | S    | Std  |
| 2.5 - 1         | 52.5                      | 9644  | -       | 2.5000   | 245.1  | PPB   | 13 Mar 2024 12:03:52 | S    | Std  |
| 5.0 - 1         | 55                        | 19885 | -       | 5.0000   | 245.1  | PPB   | 13 Mar 2024 12:06:09 | S    | Std  |
| 7.5 - 1         | 57.5                      | 29016 | -       | 7.5000   | 245.1  | PPB   | 13 Mar 2024 12:10:46 | S    | Std  |
| 10.0 - 1        | 510                       | 39318 | -       | 10.0000  | 245.1  | PPB   | 13 Mar 2024 12:23:54 | S    | Std  |
| ICV46 - 1       | ICV46                     | 15224 | 3.8837  | -245.1   |        | PPB   | 13 Mar 2024 12:32:45 | U    | SMPL |
| ICB46 - 1       | ICB46                     | -204  | -0.0584 | -245.1   |        | PPB   | 13 Mar 2024 12:35:00 | U    | SMPL |
| CCV60 - 1       | CCV60                     | 19467 | 4.9679  | -245.1   |        | PPB   | 13 Mar 2024 12:39:57 | U    | SMPL |
| CCB60 - 1       | CCB60                     | -233  | -0.0658 | -245.1   |        | PPB   | 13 Mar 2024 12:42:12 | U    | SMPL |
| CRA - 1         | CRA                       | 222   | 0.0504  | -245.1   |        | PPB   | 13 Mar 2024 12:44:29 | U    | SMPL |
| HighStd - 1     | HighStd                   | 41238 | 10.5308 | -245.1   |        | PPB   | 13 Mar 2024 12:49:21 | U    | SMPL |
| ChkStd - 1      | ChkStd                    | 27933 | 7.1311  | -245.1   |        | PPB   | 13 Mar 2024 12:51:38 | U    | SMPL |
| PB159576BL - 1  | PBW                       | -328  | -0.0901 | -245.1   |        | PPB   | 13 Mar 2024 12:53:58 | U    | SMPL |
| PB159576BS - 1  | LCSW                      | 15236 | 3.8868  | -245.1   |        | PPB   | 13 Mar 2024 12:58:36 | U    | SMPL |
| P1187-10 - 1    | LOD-MDL-WATER-04-QT1-2024 | 222   | 0.0504  | -245.1   |        | PPB   | 13 Mar 2024 13:03:30 | U    | SMPL |
| P1187-11 - 1    | LOQ-WATER-05-QT1-2024     | 881   | 0.2188  | -245.1   |        | PPB   | 13 Mar 2024 13:07:30 | U    | SMPL |
| P1675-02 - 1    | SPRING-2024               | 2652  | 0.6713  | -245.1   |        | PPB   | 13 Mar 2024 13:14:51 | U    | SMPL |
| P1675-02DUP - 1 | SPRING-2024DUP            | 2451  | 0.6200  | -245.1   |        | PPB   | 13 Mar 2024 13:17:08 | U    | SMPL |
| P1675-02MS - 1  | SPRING-2024MS             | 17305 | 4.4155  | -245.1   |        | PPB   | 13 Mar 2024 13:21:58 | U    | SMPL |
| CCV61 - 1       | CCV61                     | 20587 | 5.2541  | -245.1   |        | PPB   | 13 Mar 2024 13:24:14 | U    | SMPL |
| CCB61 - 1       | CCB61                     | -73   | -0.0250 | -245.1   |        | PPB   | 13 Mar 2024 13:27:10 | U    | SMPL |
| P1675-02MSD - 1 | SPRING-2024MSD            | 18267 | 4.6613  | -245.1   |        | PPB   | 13 Mar 2024 13:29:24 | U    | SMPL |
| P1723-02 - 1    | EFF-WW                    | -217  | -0.0617 | -245.1   |        | PPB   | 13 Mar 2024 13:31:40 | U    | SMPL |
| P1675-02LX5 - 1 |                           | 464   | 0.1123  | -245.1   |        | PPB   | 13 Mar 2024 13:36:36 | U    | SMPL |
| P1675-02A - 1   |                           | 19842 | 5.0637  | -245.1   |        | PPB   | 13 Mar 2024 13:38:52 | U    | SMPL |
| CCV62 - 1       | CCV62                     | 19184 | 4.8956  | -245.1   |        | PPB   | 13 Mar 2024 13:41:08 | U    | SMPL |
| CCB62 - 1       | CCB62                     | -263  | -0.0735 | -245.1   |        | PPB   | 13 Mar 2024 13:43:27 | U    | SMPL |