



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129866

Ambient Barometric Pressure (mmHg): 775.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	03/13/2024	11:00
2	P1734-01		1	12	100.00	>212.0	>212.0	03/13/2024	11:30
3	P1734-01DUP		1	13	100.00	>212.0	>212.0	03/13/2024	12:00
4	P1734-06		1	14	100.00	>212.0	>212.0	03/13/2024	12:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-129866

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 03/13/24 flashpoint WorkList ID : 178555 Department : Wet-Chemistry Date : 03-13-2024 10:02:10

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1734-01	COMP-1	Water	Flash Point	Cool 4 deg C	PSEG03	I31	03/12/2024	1010B
P1734-06	COMP-2	Water	Flash Point	Cool 4 deg C	PSEG03	I31	03/12/2024	1010B

Date/Time 03/13/2024 10:45
Raw Sample Received by: N.P. W.C.
Raw Sample Relinquished by: M. W. C.

Date/Time 03/13/2024 13:20
Raw Sample Received by: M. W. C.
Raw Sample Relinquished by: N.P. W.C.