



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Rubina

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB129887

Reagent/Standard	Lot/Log #
Residual chlorine ICV-LCS, 0.4PPM	WP106961
Chlorine Calibration std, 0.1ppm	WP106956
Chlorine Calibration std, 0.2ppm	WP106957
Chlorine Calibration std, 0.8ppm	WP106959
Chlorine Calibration std, 0.0ppm	WP106955
Chlorine Calibration std, 1.6ppm	WP106960
Residual Chlorine Calibration and CCV std, 0	WP106958
Total Chlorine Powder Pillows	W2759

Intercept: 0.0120

Slope: 1.0090

Regression: 0.999841

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.01		03/14/2024	12:25
2	CAL2	0.1	1	0.000	0.120	0.120	0.11	7	03/14/2024	12:28
3	CAL3	0.2	1	0.000	0.230	0.230	0.22	8	03/14/2024	12:31
4	CAL4	0.4	1	0.000	0.410	0.410	0.39	-1.5	03/14/2024	12:34
5	CAL5	0.8	1	0.000	0.810	0.810	0.79	-1.1	03/14/2024	12:37
6	CAL6	1.6	1	0.000	1.630	1.630	1.60	0.3	03/14/2024	12:40



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Rubina

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB129887

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.3900	0.3900	0.3750	03/14/2024	12:43
2	ICB				1	0.0000	0.0100	0.0100	-0.0020	03/14/2024	12:46
3	CCV1			0.4	1	0.0000	0.4200	0.4200	0.4040	03/14/2024	12:49
4	CCB1				1	0.0000	0.0000	0.0000	-0.0120	03/14/2024	12:52
5	LB129887BL	50	50		1	0.0000	0.0100	0.0100	-0.0020	03/14/2024	12:55
6	LB129887BS	50	50	0.4	1	0.0000	0.4000	0.4000	0.3850	03/14/2024	12:58
7	P1670-01	50	50		1	0.0000	0.0500	0.0500	0.0380	03/14/2024	13:01
8	P1670-01DUP	50	50		1	0.0000	0.0500	0.0500	0.0380	03/14/2024	13:04
9	P1670-01MS	50	50	0.4	1	0.0000	0.4000	0.4000	0.3850	03/14/2024	13:07
10	P1670-01MSD	50	50	0.4	1	0.0000	0.3900	0.3900	0.3750	03/14/2024	13:10
11	P1671-01	50	50		1	0.0000	0.1100	0.1100	0.0970	03/14/2024	13:13
12	P1671-02	50	50		1	0.0000	0.6200	0.6200	0.6030	03/14/2024	13:16
13	CCV2			0.4	1	0.0000	0.4100	0.4100	0.3940	03/14/2024	13:19
14	CCB2				1	0.0000	0.0000	0.0000	-0.0120	03/14/2024	13:22

WORKLIST(Hardcopy Internal Chain)

LB129887

WorkList Name : Residual Chl-031424

WorkList ID : 178580

Department : Wet-Chemistry

Date : 03-14-2024 10:56:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1670-01	GAMMA-JARINU	Water	Residual Chlorine	Cool 4 deg C	STER18	H21	02/26/2024	SM4500 Cl G
P1671-01	JS9600	Water	Residual Chlorine	Cool 4 deg C	STER18	H21	02/27/2024	SM4500 Cl G
P1671-02	JS7500	Water	Residual Chlorine	Cool 4 deg C	STER18	H21	02/27/2024	SM4500 Cl G

Date/Time 03/14/23 11:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 03/14/23 13:45
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]