



Analytical Summary Report

Analysis Method: 9041A

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129892

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/14/2024	15:30
2	CAL2	1	Water	NA	NA	20.3	6.99	03/14/2024	15:31
3	CAL3	1	Water	NA	NA	20.3	10.02	03/14/2024	15:33
4	ICV	1	Water	NA	NA	20.3	7.01	03/14/2024	15:35
5	CCV1	1	Water	NA	NA	20.3	2.01	03/14/2024	15:37
6	P1764-10	1	Solid	20.02	20	24.0	5.93	03/14/2024	15:40
7	P1764-11	1	Solid	20.01	20	25.1	5.49	03/14/2024	15:45
8	P1764-11DUP	1	Solid	20.02	20	25.2	5.50	03/14/2024	15:46
9	CCV2	1	Water	NA	NA	20.3	12.02	03/14/2024	15:50

WORKLIST(Hardcopy Internal Chain)

124892

WorkList Name : ph p1764 solids

WorkList ID : 178597

Department : Wet-Chemistry

Date : 03-14-2024 15:19:28

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1764-10	COMP-1	Solid	pH	Cool 4 deg C	DASK01	I31	03/14/2024	9041A
P1764-11	COMP-2	Solid	pH	Cool 4 deg C	DASK01	I31	03/14/2024	9041A

Date/Time 03-14-24 15:20

Raw Sample Received by: JDL

Raw Sample Relinquished by: JDL

Date/Time 03-14-24

Raw Sample Received by: JDL

Raw Sample Relinquished by: JDL