


LB129895

SFAM01.1 INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.8176e-004 slope

C= -1.2279e-001 y-intercept

Rho= 0.9996731

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g.d
0.00	0.000	-0.063	-0.063	328	0.000	328					0.0
0.05	0.050					442	0				2
0.20	0.200	0.242	0.042	2008	0.0 %	2008					2
2.50	2.500	2.454	-0.046	14179	0.0 %	14179					-2
5.00	5.000	5.180	0.180	29176	0.0 %	29176					4
7.50	7.500	7.386	-0.114	41310	0.0 %	41310					-2
10.0	10.000	10.001	0.001	55700	0.0 %	55700					0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 SSO	328	-		0 SFAM01.1	PPB	3/15/2024 8:05	Std
	0.2 S01	2008	-		0.2 SFAM01.1	PPB	3/15/2024 8:09	Std
	2.5 S02	14179	-		2.5 SFAM01.1	PPB	3/15/2024 8:11	Std
	5 S03	29176	-		5 SFAM01.1	PPB	3/15/2024 8:14	Std
	7.5 S04	41310	-		7.5 SFAM01.1	PPB	3/15/2024 8:16	Std
	10 S05	55700	-		10 SFAM01.1	PPB	3/15/2024 8:18	Std
ICV050	ICV050	20566	3.6152 -		SFAM01.1	PPB	3/15/2024 8:24	SMPL
ICB050	ICB050	353	-0.0586 -		SFAM01.1	PPB	3/15/2024 8:26	SMPL
CCV072	CCV072	29336	5.2092 -		SFAM01.1	PPB	3/15/2024 8:29	SMPL
CCB072	CCB072	281	-0.0717 -		SFAM01.1	PPB	3/15/2024 8:31	SMPL
PB159602BL	PBS602	471	-0.0372 -		SFAM01.1	PPB	3/15/2024 8:33	SMPL
P1646-01	YCML1	4502	0.6955 -		SFAM01.1	PPB	3/15/2024 8:36	SMPL
P1646-02	YCML2	5063	0.7974 -		SFAM01.1	PPB	3/15/2024 8:40	SMPL
P1646-03	YCML3	2155	0.2689 -		SFAM01.1	PPB	3/15/2024 8:42	SMPL
P1646-04	YCML4	1737	0.1929 -		SFAM01.1	PPB	3/15/2024 8:44	SMPL
P1646-05	YCML5	2844	0.3941 -		SFAM01.1	PPB	3/15/2024 8:46	SMPL
P1646-06	YCML6	2310	0.2971 -		SFAM01.1	PPB	3/15/2024 8:49	SMPL
P1646-07	YCML6D	2463	0.3249 -		SFAM01.1	PPB	3/15/2024 8:51	SMPL
P1646-08	YCML6S	15092	2.6203 -		SFAM01.1	PPB	3/15/2024 8:53	SMPL
P1646-09	YCML7	1791	0.2027 -		SFAM01.1	PPB	3/15/2024 8:55	SMPL
P1646-10	YCML8	1683	0.1831 -		SFAM01.1	PPB	3/15/2024 8:58	SMPL
P1646-11	YCML9	2592	0.3483 -		SFAM01.1	PPB	3/15/2024 9:00	SMPL
P1646-12	YCMM0	2441	0.3209 -		SFAM01.1	PPB	3/15/2024 9:02	SMPL
P1646-13	YCMM1	2325	0.2998 -		SFAM01.1	PPB	3/15/2024 9:04	SMPL
PB159603BL	PBS603	329	-0.063 -		SFAM01.1	PPB	3/15/2024 9:07	SMPL
P1761-01	E0731	1138	0.084 -		SFAM01.1	PPB	3/15/2024 9:09	SMPL
P1761-02	E0732	556	-0.0217 -		SFAM01.1	PPB	3/15/2024 9:11	SMPL
P1761-03	E0733	435	-0.0437 -		SFAM01.1	PPB	3/15/2024 9:14	SMPL
P1761-04	E0734	509	-0.0303 -		SFAM01.1	PPB	3/15/2024 9:17	SMPL
P1761-05	E0735	133	-0.0986 -		SFAM01.1	PPB	3/15/2024 9:19	SMPL
CCV073	CCV073	29677	5.2712 -		SFAM01.1	PPB	3/15/2024 9:22	SMPL
CCB073	CCB073	353	-0.0586 -		SFAM01.1	PPB	3/15/2024 9:24	SMPL
P1761-06	E0736	3540	0.5206 -		SFAM01.1	PPB	3/15/2024 9:26	SMPL
P1761-07	E0737	671	-0.0008 -		SFAM01.1	PPB	3/15/2024 9:28	SMPL
P1761-08	E0738	1054	0.0688 -		SFAM01.1	PPB	3/15/2024 9:31	SMPL
P1761-09	E0739	136	-0.0981 -		SFAM01.1	PPB	3/15/2024 9:36	SMPL
P1761-10	E0740	270	-0.0737 -		SFAM01.1	PPB	3/15/2024 9:38	SMPL
P1761-12	E0756	8204	1.3683 -		SFAM01.1	PPB	3/15/2024 9:40	SMPL
P1761-13	E0757	409	-0.0485 -		SFAM01.1	PPB	3/15/2024 9:43	SMPL
P1761-14	E0758	394	-0.0512 -		SFAM01.1	PPB	3/15/2024 9:45	SMPL
P1761-15	E0759	2677	0.3638 -		SFAM01.1	PPB	3/15/2024 9:47	SMPL
P1761-16	E0760	395	-0.051 -		SFAM01.1	PPB	3/15/2024 9:49	SMPL
P1761-17	E0761	266	-0.0744 -		SFAM01.1	PPB	3/15/2024 9:52	SMPL
P1761-18	E0761D	463	-0.0386 -		SFAM01.1	PPB	3/15/2024 9:55	SMPL
P1761-19	E0761S	14275	2.4718 -		SFAM01.1	PPB	3/15/2024 9:57	SMPL
CCV074	CCV074	29796	5.2928 -		SFAM01.1	PPB	3/15/2024 9:59	SMPL

LB129895

INSTRUMENT ID : CV1

CCB074

CCB074

347 -0.0597 -

SFAM01.1 PPB

3/15/2024 10:01 SMPL