



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129898

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/15/2024	09:15
2	CAL2	1	Water	NA	NA	20.3	7.00	03/15/2024	09:16
3	CAL3	1	Water	NA	NA	20.3	10.02	03/15/2024	09:19
4	ICV	1	Water	NA	NA	20.4	7.00	03/15/2024	09:20
5	CCV1	1	Water	NA	NA	20.3	2.01	03/15/2024	09:25
6	P1734-06	1	Water	NA	NA	20.4	7.23	03/15/2024	09:30
7	P1772-12	1	Water	NA	NA	21.2	8.41	03/15/2024	09:40
8	P1772-12DUP	1	Water	NA	NA	21.3	8.42	03/15/2024	09:41
9	CCV2	1	Water	NA	NA	20.3	12.02	03/15/2024	09:42

UP 129898

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH P1772 WorkList ID : 178617 Department : Wet-Chemistry Date : 03-15-2024 09:09:35

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1734-06	17 COMP-2	Water	pH	Cool 4 deg C	PSEG03	I31	03/12/2024	9040C
P1772-12	17 1133	Water	pH	Cool 4 deg C	PSEG03	O21	03/14/2024	9040C

Date/Time 03/15/24 09:10
Raw Sample Received by: JD (signature)
Raw Sample Relinquished by: JD (signature)

Date/Time 03/15/24 11:00
Raw Sample Received by: JD (signature)
Raw Sample Relinquished by: JD (signature)