



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129912

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	03/15/2024	10:45
2	P1772-06		1	12	100.00	>212.0	>212.0	03/15/2024	11:15
3	P1772-06DUP		1	13	100.00	>212.0	>212.0	03/15/2024	11:45
4	P1772-07		1	14	100.00	>212.0	>212.0	03/15/2024	12:15
5	P1772-08		1	12	100.00	>212.0	>212.0	03/15/2024	12:45
6	P1772-09		1	13	100.00	>212.0	>212.0	03/15/2024	13:15
7	P1772-10		1	14	100.00	>212.0	>212.0	03/15/2024	13:45
8	P1772-11		1	12	100.00	>212.0	>212.0	03/15/2024	14:15
9	P1772-12		1	14	100.00	>212.0	>212.0	03/15/2024	14:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain) LB-129912

WorkList Name : 03/15/24 flashpoint WorkList ID : 178613 Department : Wet-Chemistry Date : 03-15-2024 08:23:51

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1772-06	1350	Water	Flash Point	Cool 4 deg C	PSEG03	O21	03/14/2024	1010B
P1772-07	1376	Water	Flash Point	Cool 4 deg C	PSEG03	O21	03/14/2024	1010B
P1772-08	1346	Water	Flash Point	Cool 4 deg C	PSEG03	O21	03/14/2024	1010B
P1772-09	1351	Water	Flash Point	Cool 4 deg C	PSEG03	O21	03/14/2024	1010B
P1772-10	1352	Water	Flash Point	Cool 4 deg C	PSEG03	O21	03/14/2024	1010B
P1772-11	1349	Water	Flash Point	Cool 4 deg C	PSEG03	O21	03/14/2024	1010B
P1772-12	1133	Water	Flash Point	Cool 4 deg C	PSEG03	O21	03/14/2024	1010B

Date/Time 03/15/2024 10:30
Raw Sample Received by: N.P.W.C
Raw Sample Relinquished by: [Signature]

Date/Time 03/15/2024 15:15
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: N.P.W.C