



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129922

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/18/2024	12:40
2	CAL2	1	Water	NA	NA	20.3	7.02	03/18/2024	12:41
3	CAL3	1	Water	NA	NA	20.4	10.02	03/18/2024	12:42
4	ICV	1	Water	NA	NA	20.3	6.98	03/18/2024	12:45
5	CCV1	1	Water	NA	NA	20.2	2.01	03/18/2024	12:48
6	P1775-01	1	Water	NA	NA	20.1	6.66	03/18/2024	12:55
7	P1775-01DUP	1	Water	NA	NA	20.2	6.67	03/18/2024	13:00
8	CCV2	1	Water	NA	NA	20.3	12.02	03/18/2024	13:05