



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129923

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	03/18/2024	11:15
2	P1780-01		1	14	100.00	>212.0	>212.0	03/18/2024	11:45
3	P1780-01DUP		1	13	100.00	>212.0	>212.0	03/18/2024	12:15
4	P1781-01		1	12	100.00	>212.0	>212.0	03/18/2024	12:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-129923

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 3/18/24 flashpoint

WorkList ID : 178653

Department : Wet-Chemistry

Date : 03-18-2024 08:06:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1780-01	2011	Water	Flash Point	Cool 4 deg C	PSEG03	I41	03/15/2024	1010B
P1781-01	1372	Water	Flash Point	Cool 4 deg C	PSEG03	I41	03/15/2024	1010B

Date/Time 03/18/2024 11:00
Raw Sample Received by: N.P.W.C.
Raw Sample Relinquished by: N.P.W.C.

Date/Time 03/18/2024 13:35
Raw Sample Received by: N.P.W.C.
Raw Sample Relinquished by: N.P.W.C.