



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB129931

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/18/2024	14:05
2	CAL2	1	Water	NA	NA	20.2	7.00	03/18/2024	14:10
3	CAL3	1	Water	NA	NA	20.3	10.02	03/18/2024	14:15
4	ICV	1	Water	NA	NA	20.3	7.01	03/18/2024	14:20
5	CCV1	1	Water	NA	NA	20.3	2.01	03/18/2024	14:25
6	P1790-01	1	Water	NA	NA	20.2	6.21	03/18/2024	14:30
7	P1790-01DUP	1	Water	NA	NA	20.3	6.22	03/18/2024	14:31
8	P1790-02	1	Water	NA	NA	20.3	6.60	03/18/2024	14:40
9	P1790-03	1	Water	NA	NA	20.6	6.32	03/18/2024	14:45
10	P1790-04	1	Water	NA	NA	20.4	6.26	03/18/2024	14:50
11	P1790-05	1	Water	NA	NA	20.3	6.30	03/18/2024	15:00
12	P1790-06	1	Water	NA	NA	20.4	6.67	03/18/2024	15:10
13	CCV2	1	Water	NA	NA	20.3	12.02	03/18/2024	15:15

UP 129931

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH P1790		WorkList ID : 178658		Department : Wet-Chemistry		Date : 03-18-2024 13:43:33		
Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1790-01	COMP-1	Water	pH	Cool 4 deg C	PSEG03	I21	03/18/2024	9040C
P1790-02	NWB-2025	Water	pH	Cool 4 deg C	PSEG03	I21	03/18/2024	9040C
P1790-03	NWB-2026	Water	pH	Cool 4 deg C	PSEG03	I21	03/18/2024	9040C
P1790-04	NWB-2027	Water	pH	Cool 4 deg C	PSEG03	I21	03/18/2024	9040C
P1790-05	NWB-2028	Water	pH	Cool 4 deg C	PSEG03	I21	03/18/2024	9040C
P1790-06	NWB-2029	Water	pH	Cool 4 deg C	PSEG03	I21	03/18/2024	9040C

Date/Time	03/18/24	14100	Date/Time	03/18/24	17100
Raw Sample Received by:	SB CEC		Raw Sample Received by:	J.C. (sm)	
Raw Sample Relinquished by:	J.C. (sm)		Raw Sample Relinquished by:	SB CEC	