

LB1299410

7470A

INSTRUMENT ID: CV1

 μ Abs.:

47341

Linear

A= 0.0000e+000

B= 2.0704e-004

C= -9.8766e-002

Rho= 0.9982320

Accept=Accepted

Max. Conc:—

10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/1, Δ
0.0	0.000	-0.036	-0.036	304	0.000	304	0				—
0.2	0.200	0.203	0.003	1457	0.0 %	1457					2
2.5	2.500	2.517	0.017	12635	0.0 %	12635					1
5.0	5.000	4.883	-0.117	24061	0.0 %	24061					— 2
7.5	7.500	7.930	0.430	38779	0.0 %	38779					6
10.0	10.000	9.703	-0.297	47341	0.0 %	47341					— 3

LB129940 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Mar 2024 12:05:31

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	304	-	0.00007470A	PPB	19 Mar 2024 12:17:45	S	Std
0.2 - 1	50-2	1457	-	0.20007470A	PPB	19 Mar 2024 12:20:02	S	Std
2.5 - 1	52.5	12635	-	2.50007470A	PPB	19 Mar 2024 12:22:19	S	Std
5.0 - 1	55	24061	-	5.00007470A	PPB	19 Mar 2024 12:24:36	S	Std
7.5 - 1	57.5	38779	-	7.50007470A	PPB	19 Mar 2024 12:26:55	S	Std
10.0 - 1	51	47341	-	10.00007470A	PPB	19 Mar 2024 12:31:58	S	Std
ICV53 - 1	ICV53	18235	3.6766	-7470A	PPB	19 Mar 2024 12:37:36	U	SMPL
ICB53 - 1	ICB53	599	0.0253	-7470A	PPB	19 Mar 2024 12:39:50	U	SMPL
CCV81 - 1	CCV81	23322	4.7299	-7470A	PPB	19 Mar 2024 12:42:08	U	SMPL
CCB81 - 1	CCB81	583	0.0219	-7470A	PPB	19 Mar 2024 12:44:23	U	SMPL
CRA - 1	CRA	1428	0.1969	-7470A	PPB	19 Mar 2024 12:46:41	U	SMPL
HighStd - 1	HighStd	47727	9.7827	-7470A	PPB	19 Mar 2024 12:48:55	U	SMPL
ChkStd - 1	ChkStd	33014	6.7365	-7470A	PPB	19 Mar 2024 12:51:11	U	SMPL
PB159624BL - 1	PBW	653	0.0364	-7470A	PPB	19 Mar 2024 12:56:07	U	SMPL
PB159624BS - 1	LCSW	18042	3.6367	-7470A	PPB	19 Mar 2024 12:58:22	U	SMPL
P1746-05 - 1	SB-01-COMP	134	-0.0710	-7470A	PPB	19 Mar 2024 13:00:39	U	SMPL
P1746-07 - 1	SB-02-COMP	271	-0.0427	-7470A	PPB	19 Mar 2024 13:02:58	U	SMPL
P1746-10 - 1	SB-03-COMP	250	-0.0470	-7470A	PPB	19 Mar 2024 13:05:14	U	SMPL
P1746-14 - 1	SB-04-COMP	212	-0.0549	-7470A	PPB	19 Mar 2024 13:07:31	U	SMPL
P1746-18 - 1	SB-05-COMP	283	-0.0402	-7470A	PPB	19 Mar 2024 13:09:46	U	SMPL
CCV82 - 1	CCV82	25896	5.2628	-7470A	PPB	19 Mar 2024 13:12:02	U	SMPL
CCB82 - 1	CCB82	601	0.0257	-7470A	PPB	19 Mar 2024 13:14:17	U	SMPL
P1746-20 - 1	SB-06-COMP	224	-0.0524	-7470A	PPB	19 Mar 2024 13:16:35	U	SMPL
P1771-02 - 1	COMP-1	272	-0.0425	-7470A	PPB	19 Mar 2024 13:18:50	U	SMPL
P1771-04 - 1	COMP-2	217	-0.0538	-7470A	PPB	19 Mar 2024 13:21:05	U	SMPL
P1771-06 - 1	COMP-3	243	-0.0485	-7470A	PPB	19 Mar 2024 13:23:21	U	SMPL
P1771-08 - 1	COMP-4	315	-0.0335	-7470A	PPB	19 Mar 2024 13:25:37	U	SMPL
P1771-08DUP - 1	COMP-4DUP	463	-0.0029	-7470A	PPB	19 Mar 2024 13:27:53	U	SMPL
P1771-08MS - 1	COMP-4MS	21621	4.3777	-7470A	PPB	19 Mar 2024 13:32:56	U	SMPL
P1771-08MSD - 1	COMP-4MSD	21593	4.3719	-7470A	PPB	19 Mar 2024 13:35:12	U	SMPL
PB159659BL - 1	PBW	171	-0.0634	-7470A	PPB	19 Mar 2024 13:37:31	U	SMPL
PB159659BS - 1	LCSW	18008	3.6296	-7470A	PPB	19 Mar 2024 13:39:50	U	SMPL
CCV83 - 1	CCV83	27002	5.4918	-7470A	PPB	19 Mar 2024 13:42:05	U	SMPL
CCB83 - 1	CCB83	490	0.0027	-7470A	PPB	19 Mar 2024 13:44:23	U	SMPL
P1747-01 - 1	MW-01	248	-0.0474	-7470A	PPB	19 Mar 2024 13:46:42	U	SMPL
P1747-02 - 1	MW-01-DUP	307	-0.0352	-7470A	PPB	19 Mar 2024 13:48:58	U	SMPL
P1747-04 - 1	MW-02	400	-0.0159	-7470A	PPB	19 Mar 2024 13:51:13	U	SMPL
P1747-05 - 1	MW-04	431	-0.0095	-7470A	PPB	19 Mar 2024 13:53:29	U	SMPL
P1747-07 - 1	MW-01	223	-0.0526	-7470A	PPB	19 Mar 2024 13:55:45	U	SMPL
P1747-08 - 1	MW-01-DUP	333	-0.0298	-7470A	PPB	19 Mar 2024 13:58:01	U	SMPL
P1747-09 - 1	MW-02	252	-0.0466	-7470A	PPB	19 Mar 2024 14:00:17	U	SMPL
P1747-10 - 1	MW-04	361	-0.0240	-7470A	PPB	19 Mar 2024 14:02:34	U	SMPL
P1769-01 - 1	MW-2	1853	0.2849	-7470A	PPB	19 Mar 2024 14:04:51	U	SMPL
P1769-02 - 1	MW-1	391	-0.0178	-7470A	PPB	19 Mar 2024 14:07:06	U	SMPL
CCV84 - 1	CCV84	26131	5.3114	-7470A	PPB	19 Mar 2024 14:09:23	U	SMPL
CCB84 - 1	CCB84	437	-0.0083	-7470A	PPB	19 Mar 2024 14:11:39	U	SMPL
P1769-03 - 1	FB-3-14-24	246	-0.0478	-7470A	PPB	19 Mar 2024 14:13:57	U	SMPL
P1772-06 - 1	1350	1780	0.2698	-7470A	PPB	19 Mar 2024 14:16:13	U	SMPL
P1772-08 - 1	1346	2004	0.3161	-7470A	PPB	19 Mar 2024 14:23:00	U	SMPL
P1772-08DUP - 1	1346DUP	2226	0.3621	-7470A	PPB	19 Mar 2024 14:27:53	U	SMPL
P1772-08MS - 1	1346MS	19131	3.8621	-7470A	PPB	19 Mar 2024 14:32:49	U	SMPL
P1772-08MSD - 1	1346MSD	18440	3.7191	-7470A	PPB	19 Mar 2024 14:35:04	U	SMPL
CCV85 - 1	CCV85	25254	5.1299	-7470A	PPB	19 Mar 2024 14:40:07	U	SMPL
CCB85 - 1	CCB85	840	0.0751	-7470A	PPB	19 Mar 2024 14:42:23	U	SMPL
PB159589TB - 1	PB159589TB	389	-0.0182	-7470A	PPB	19 Mar 2024 14:44:43	U	SMPL
P1771-08LX5 - 1		42	-0.0901	-7470A	PPB	19 Mar 2024 14:47:37	U	SMPL
P1771-08A - 1		21197	4.2899	-7470A	PPB	19 Mar 2024 14:49:52	U	SMPL
P1772-08LX5 - 1		621	0.0298	-7470A	PPB	19 Mar 2024 14:52:08	U	SMPL
P1772-08A - 1		22422	4.5435	-7470A	PPB	19 Mar 2024 14:54:26	U	SMPL
CCV86 - 1	CCV86	26097	5.3044	-7470A	PPB	19 Mar 2024 14:57:22	U	SMPL
CCB86 - 1	CCB86	733	0.0530	-7470A	PPB	19 Mar 2024 14:59:37	U	SMPL