



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129943

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	03/19/2024	11:00
2	P1788-01		1	12	100.00	>212.0	>212.0	03/19/2024	11:30
3	P1788-01DUP		1	13	100.00	>212.0	>212.0	03/19/2024	12:00
4	P1788-02		1	14	100.00	>212.0	>212.0	03/19/2024	12:30
5	P1790-01		1	12	100.00	>212.0	>212.0	03/19/2024	13:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

L8.129943

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 03/19/24 flashpoint w WorkList ID : 178693 Department : Wet-Chemistry Date : 03-19-2024 12:05:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1788-01	PEZOMETER-200	Water	Flash Point	Cool 4 deg C	MIRM01	O21	03/18/2024	1010B
P1788-02	PEZOMETER-206	Water	Flash Point	Cool 4 deg C	MIRM01	O21	03/18/2024	1010B
P1790-01	COMP-1	Water	Flash Point	Cool 4 deg C	PSEG03	I21	03/18/2024	1010B

Date/Time 03/19/2024 10:45
Raw Sample Received by: N.P. W.C.
Raw Sample Relinquished by: [Signature]

Date/Time 03/19/2024 14:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: N.P. W.C.