

Analysis Method: 9034

ANALYST: Nikita

Parameter: Reactive Sulfide

SUPERVISOR REVIEW BY: Iwona

Run Number: LB129945

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W2976
IODINE SOLUTION .025N 1L	W3050
Starch Solution, 4L	W2977

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB159674BL		1	5.00	50	2.00	0.00	1.94	1.94	0.06	0.00	0.00	03/19/2024	15:15
2	P1616-02		1	5.02	50	2.00	0.00	1.86	1.86	0.14	0.08	6.37	03/19/2024	15:18
3	P1616-02DUP		1	5.05	50	2.00	0.00	1.86	1.86	0.14	0.08	6.34	03/19/2024	15:21
4	P1616-04		1	5.01	50	2.00	0.00	1.90	1.90	0.10	0.04	3.19	03/19/2024	15:24
5	P1786-03		1	5.03	50	2.00	0.00	1.90	1.90	0.10	0.04	3.18	03/19/2024	15:27
6	P1792-02		1	5.04	50	2.00	0.00	1.92	1.92	0.08	0.02	1.59	03/19/2024	15:30
7	P1793-02		1	5.04	50	2.00	0.00	1.93	1.93	0.07	0.01	0.79	03/19/2024	15:33

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume