



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB129950

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/20/2024	08:05
2	CAL2	1	Water	NA	NA	20.3	6.99	03/20/2024	08:06
3	CAL3	1	Water	NA	NA	20.3	10.03	03/20/2024	08:09
4	ICV	1	Water	NA	NA	20.6	7.01	03/20/2024	08:10
5	CCV1	1	Water	NA	NA	20.4	2.02	03/20/2024	08:50
6	P1799-01	1	Water	NA	NA	21.7	6.17	03/20/2024	09:00
7	P1799-02	1	Water	NA	NA	21.6	5.86	03/20/2024	09:05
8	P1799-03	1	Water	NA	NA	21.6	5.54	03/20/2024	09:14
9	P1799-04	1	Water	NA	NA	21.7	6.17	03/20/2024	09:20
10	P1800-01	1	Water	NA	NA	21.5	7.24	03/20/2024	09:27
11	P1807-05	1	Water	NA	NA	21.8	7.33	03/20/2024	09:35
12	P1807-05DUP	1	Water	NA	NA	21.9	7.34	03/20/2024	09:36
13	CCV2	1	Water	NA	NA	20.4	12.02	03/20/2024	09:40

WORKLIST(Hardcopy Internal Chain)

VB 120050

WorkList Name : ph p1800

WorkList ID : 178700

Department : Wet-Chemistry

Date : 03-20-2024 07:48:10

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1799-01	C 50295	Water	pH	Cool 4 deg C	PSEG03	O31	03/19/2024	9040C
P1799-02	C 50296	Water	pH	Cool 4 deg C	PSEG03	O31	03/19/2024	9040C
P1799-03	C 50297	Water	pH	Cool 4 deg C	PSEG03	O31	03/19/2024	9040C
P1799-04	C 50298	Water	pH	Cool 4 deg C	PSEG03	O31	03/19/2024	9040C
P1800-01	C 1394	Water	pH	Cool 4 deg C	PSEG03	O31	03/19/2024	9040C
P1807-05	C 3462	Water	pH	Cool 4 deg C	PSEG03	O31	03/19/2024	9040C

Date/Time 03/20/24 08:00

Raw Sample Received by: JB (COC)

Raw Sample Relinquished by: JB SR

Date/Time 03/20/24 12:30

Raw Sample Received by: JB SR

Raw Sample Relinquished by: JB (COC)