



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129952

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	03/20/2024	08:30
2	P1772-04		1	8	50.00	122.0	122.0	03/20/2024	09:00
3	P1772-04DUP		1	8	50.00	122.0	122.0	03/20/2024	09:30
4	P1772-05		1	8	25.00	77.0	77.0	03/20/2024	10:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-129952

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 03/19/24 flashpoint

WorkList ID : 178675

Department : Wet-Chemistry

Date : 03-19-2024 08:18:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P-1772-04	1313	Solid	Flash Point	Cool 4 deg C	PSEG03	O21	03/14/2024	1010B
P-1772-05	1280	Solid	Flash Point	Cool 4 deg C	PSEG03	O21	03/14/2024	1010B

Date/Time03/19/20248:15

Raw Sample Received by: N.P. W.C

Raw Sample Relinquished by: M GEE

Date/Time03/19/202410:50

Raw Sample Received by: M GEE

Raw Sample Relinquished by: N.P. W.C