



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129961

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	03/20/2024	15:30
2	P1807-01		1	8	40.00	104.0	103.0	03/20/2024	16:00
3	P1807-01DUP		1	8	40.00	104.0	103.0	03/20/2024	16:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB. 129961

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 03/20/24 flashpoint WorkList ID : 178730 Department : Wet-Chemistry Date : 03-20-2024 14:03:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1807-01	3461	Solid	Flash Point	Cool 4 deg C	PSEG03	O31	03/19/2024	1010B

Date/Time 03/20/2024 15:20
Raw Sample Received by: N.P. W.C
Raw Sample Relinquished by: AB1681

Date/Time 03/20/2024 16:15
Raw Sample Received by: W.C. QC1
Raw Sample Relinquished by: N.P. W.C