

LB129967


SFAM01.1

INSTRUMENT 10? CV2

Linear

 μ Abs.:

36322



A= 0.0000e+000

B= 2.7686e-004 slope

C= -7.6495e-002 $\rightarrow$ intercept

Rho= 0.9998132

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.056	0.056	480	0.000	480					
0.05	0.050					442	0				=
0.20	0.200	0.232	0.032	1115	0.0 %	1115					16
2.50	2.500	2.351	-0.149	8768	0.0 %	8768					-6
5.00	5.000	5.027	0.027	18432	0.0 %	18432					1
7.50	7.500	7.554	0.054	27563	0.0 %	27563					1
10.0	10.000	9.979	-0.021	36322	0.0 %	36322					0

LB129967

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	480	-		0 SFAM01.1	PPB	3/20/2024 16:43	Std
	0.2 S01	1115	-		0.2 SFAM01.1	PPB	3/20/2024 16:46	Std
	2.5 S02	8768	-		2.5 SFAM01.1	PPB	3/20/2024 16:48	Std
	5 S03	18432	-		5 SFAM01.1	PPB	3/20/2024 16:50	Std
	7.5 S04	27563	-		7.5 SFAM01.1	PPB	3/20/2024 16:53	Std
	10 S05	36322	-		10 SFAM01.1	PPB	3/20/2024 16:55	Std
ICV056	ICV056	15519	4.22 -		SFAM01.1	PPB	3/20/2024 17:01	SMPL
ICB056	ICB056	403	0.0351 -		SFAM01.1	PPB	3/20/2024 17:03	SMPL
CCV092	CCV092	18934	5.1655 -		SFAM01.1	PPB	3/20/2024 17:05	SMPL
CCB092	CCB092	485	0.0578 -		SFAM01.1	PPB	3/20/2024 17:08	SMPL
PB159697BL	PBS697	637	0.0999 -		SFAM01.1	PPB	3/20/2024 17:10	SMPL
P1794-01	E0764	2694	0.6694 -		SFAM01.1	PPB	3/20/2024 17:12	SMPL
P1794-02	E0765	6012	1.588 -		SFAM01.1	PPB	3/20/2024 17:16	SMPL
P1794-03	E0766	3484	0.8881 -		SFAM01.1	PPB	3/20/2024 17:18	SMPL
P1794-04	E0767	2804	0.6998 -		SFAM01.1	PPB	3/20/2024 17:20	SMPL
P1794-05	E0773	8782	2.3549 -		SFAM01.1	PPB	3/20/2024 17:30	SMPL
P1794-06	E0773D	9545	2.5661 -		SFAM01.1	PPB	3/20/2024 17:33	SMPL
P1794-07	E0773S	19205	5.2405 -		SFAM01.1	PPB	3/20/2024 17:35	SMPL
CCV093	CCV093	20479	5.5932 -		SFAM01.1	PPB	3/20/2024 17:40	SMPL
CCB093	CCB093	492	0.0597 -		SFAM01.1	PPB	3/20/2024 17:42	SMPL