



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129973

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/21/2024	08:10
2	CAL2	1	Water	NA	NA	20.3	6.99	03/21/2024	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	03/21/2024	08:15
4	ICV	1	Water	NA	NA	20.5	7.01	03/21/2024	08:20
5	CCV1	1	Water	NA	NA	20.5	2.01	03/21/2024	08:25
6	P1820-01	1	Water	NA	NA	20.6	6.30	03/21/2024	08:35
7	P1820-01DUP	1	Water	NA	NA	20.7	6.31	03/21/2024	08:36
8	P1821-02	1	Water	NA	NA	20.2	9.17	03/21/2024	08:45
9	P1821-03	1	Water	NA	NA	20.3	8.49	03/21/2024	08:50
10	CCV2	1	Water	NA	NA	20.5	12.02	03/21/2024	09:00