

LB129976

7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38607

A= 0.0000e+000

B= 2.5842e-004

C= -4.5624e-002

Rho= 0.9998651

Accept=Accepted

Max. Conc: — 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Avg
0.0	0.000	0.019	0.019	249	0.000	249	0				1
0.2	0.200	0.202	0.002	959	0.0 %	959					2
2.5	2.500	2.447	-0.053	9647	0.0 %	9647					2
5.0	5.000	4.983	-0.017	19458	0.0 %	19458					0
7.5	7.500	7.618	0.118	29655	0.0 %	29655					2
10.0	10.000	9.931	-0.069	38607	0.0 %	38607					1

LB129976 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Mar 2024 10:40:03

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	249	-	0.0000	7470A	PPB	21 Mar 2024 10:46:00	S	Std
0.2 - 1	50.2	959	-	0.2000	7470A	PPB	21 Mar 2024 10:48:17	S	Std
2.5 - 1	52.5	9647	-	2.5000	7470A	PPB	21 Mar 2024 10:50:35	S	Std
5.0 - 1	55	19458	-	5.0000	7470A	PPB	21 Mar 2024 10:52:51	S	Std
7.5 - 1	57.5	29655	-	7.5000	7470A	PPB	21 Mar 2024 10:55:10	S	Std
10.0 - 1	5N	38607	-	10.0000	7470A	PPB	21 Mar 2024 11:05:27	S	Std
ICV57 - 1	ICV57	17131	4.3814	-	7470A	PPB	21 Mar 2024 11:08:27	U	SMPL
ICB57 - 1	ICB57	116	-0.0156	-	7470A	PPB	21 Mar 2024 11:13:20	U	SMPL
CCV94 - 1	CCV94	19301	4.9421	-	7470A	PPB	21 Mar 2024 11:15:34	U	SMPL
CCB94 - 1	CCB94	18	-0.0410	-	7470A	PPB	21 Mar 2024 11:17:49	U	SMPL
CRA - 1	CRA	914	0.1906	-	7470A	PPB	21 Mar 2024 11:20:07	U	SMPL
HighStd - 1	HighStd	38534	9.9123	-	7470A	PPB	21 Mar 2024 11:22:22	U	SMPL
ChkStd - 1	ChkStd	26406	6.7782	-	7470A	PPB	21 Mar 2024 11:24:37	U	SMPL
PB159670BL - 1	PBW	-135	-0.0805	-	7470A	PPB	21 Mar 2024 11:26:59	U	SMPL
PB159670BS - 1	LCSW	15496	3.9589	-	7470A	PPB	21 Mar 2024 11:29:19	U	SMPL
P1616-02 - 1	196-BAY-AVENUE	-25	-0.0521	-	7470A	PPB	21 Mar 2024 11:31:36	U	SMPL
P1616-04 - 1	146-BAY-AVENUE	177	0.0001	-	7470A	PPB	21 Mar 2024 11:33:55	U	SMPL
P1786-03 - 1	WASTE-COMPOSITE	74	-0.0265	-	7470A	PPB	21 Mar 2024 11:36:12	U	SMPL
P1792-02 - 1	72-11977	387	0.0544	-	7470A	PPB	21 Mar 2024 11:38:28	U	SMPL
P1792-02DUP - 1	72-11977DUP	556	0.0981	-	7470A	PPB	21 Mar 2024 11:40:43	U	SMPL
CCV95 - 1	CCV95	18975	4.8579	-	7470A	PPB	21 Mar 2024 11:42:59	U	SMPL
CCB95 - 1	CCB95	125	-0.0133	-	7470A	PPB	21 Mar 2024 11:45:14	U	SMPL
P1792-02MS - 1	72-11977MS	16331	4.1746	-	7470A	PPB	21 Mar 2024 11:47:31	U	SMPL
P1792-02MSD - 1	72-11977MSD	16403	4.1832	-	7470A	PPB	21 Mar 2024 11:49:47	U	SMPL
PB159648TB - 1	PB159648TB	-43	-0.0567	-	7470A	PPB	21 Mar 2024 11:52:04	U	SMPL
P1792-02LX5 - 1		41	-0.0350	-	7470A	PPB	21 Mar 2024 11:54:23	U	SMPL
P1792-02A - 1		16205	4.1421	-	7470A	PPB	21 Mar 2024 11:58:20	U	SMPL
CCV96 - 1	CCV96	19315	4.9458	-	7470A	PPB	21 Mar 2024 12:00:36	U	SMPL
CCB96 - 1	CCB96	-17	-0.0500	-	7470A	PPB	21 Mar 2024 12:02:55	U	SMPL