


LB129978

7471B

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 1.5134e-004

C= -1.4601e-002

Rho= 0.9989256

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%1)
0.0	0.000	0.042	0.042	374	0.000	374					-
0.2	0.200	0.177	-0.023	1265	0.0 %	1265					-12
2.5	2.500	2.646	0.146	17582	0.0 %	17582					6
5.0	5.000	4.936	-0.064	32712	0.0 %	32712					-1
7.5	7.500	7.183	-0.317	47557	0.0 %	47557					-4
10.0	10.000	10.217	0.217	67605	0.0 %	67605					2

LB129978 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Mar 2024 12:39:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	51	374	-	0.0000	7471B	PPB	21 Mar 2024 13:08:38	S	Std
0.2 - 1	51.2	1265	-	0.2000	7471B	PPB	21 Mar 2024 13:13:04	S	Std
2.5 - 1	52.5	17582	-	2.5000	7471B	PPB	21 Mar 2024 13:15:22	S	Std
5.0 - 1	55	32712	-	5.0000	7471B	PPB	21 Mar 2024 13:17:39	S	Std
7.5 - 1	57.5	47557	-	7.5000	7471B	PPB	21 Mar 2024 13:22:42	S	Std
10.0 - 1	510	67605	-	10.0000	7471B	PPB	21 Mar 2024 13:27:55	S	Std
ICV58 - 1	ICV58	25278	3.8109	-	7471B	PPB	21 Mar 2024 13:30:49	U	SMPL
ICB58 - 1	ICB58	481	0.0582	-	7471B	PPB	21 Mar 2024 13:35:41	U	SMPL
CCV97 - 1	CCV97	32962	4.9738	-	7471B	PPB	21 Mar 2024 13:37:56	U	SMPL
CCB97 - 1	CCB97	198	0.0154	-	7471B	PPB	21 Mar 2024 13:40:11	U	SMPL
CRA - 1	CRA	1314	0.1843	-	7471B	PPB	21 Mar 2024 13:42:31	U	SMPL
HighStd - 1	HighStd	67896	10.2606	-	7471B	PPB	21 Mar 2024 13:44:46	U	SMPL
ChkStd - 1	ChkStd	45385	6.8538	-	7471B	PPB	21 Mar 2024 13:47:02	U	SMPL
PB159698BL - 1	PBS	463	0.0555	-	7471B	PPB	21 Mar 2024 13:52:21	U	SMPL
PB159698BS - 1	LCSS	26942	4.0627	-	7471B	PPB	21 Mar 2024 13:54:36	U	SMPL
P1616-01 - 1	196-BAY-AVENUE	2109	0.3046	-	7471B	PPB	21 Mar 2024 13:56:52	U	SMPL
P1616-03 - 1	146-BAY-AVENUE	3908	0.5768	-	7471B	PPB	21 Mar 2024 13:59:13	U	SMPL
P1786-01 - 1	TP-15	1695	0.2419	-	7471B	PPB	21 Mar 2024 14:01:31	U	SMPL
P1786-02 - 1	TP-22	2797	0.4087	-	7471B	PPB	21 Mar 2024 14:03:49	U	SMPL
P1786-02DUP - 1	TP-22DUP	3163	0.4641	-	7471B	PPB	21 Mar 2024 14:06:04	U	SMPL
CCV98 - 1	CCV98	33015	4.9818	-	7471B	PPB	21 Mar 2024 14:08:20	U	SMPL
CCB98 - 1	CCB98	752	0.0992	-	7471B	PPB	21 Mar 2024 14:10:37	U	SMPL
P1786-02MS - 1	TP-22MS	29934	4.5155	-	7471B	PPB	21 Mar 2024 14:12:59	U	SMPL
P1786-02MSD - 1	TP-22MSD	29721	4.4833	-	7471B	PPB	21 Mar 2024 14:15:15	U	SMPL
P1791-01 - 1	SU-04-03182024	3638	0.5360	-	7471B	PPB	21 Mar 2024 14:17:35	U	SMPL
P1792-01 - 1	72-11977	5301	0.7876	-	7471B	PPB	21 Mar 2024 14:19:56	U	SMPL
P1793-02 - 1	OILY-DEBRIS-DRUMS	1041	0.1429	-	7471B	PPB	21 Mar 2024 14:22:13	U	SMPL
P1801-01 - 1	OR-03-031924	2730	0.3985	-	7471B	PPB	21 Mar 2024 14:24:31	U	SMPL
P1803-01 - 1	TRENCH-PIT	3167	0.4647	-	7471B	PPB	21 Mar 2024 14:26:48	U	SMPL
P1804-01 - 1	JC-03-03192024-A	7738	1.1564	-	7471B	PPB	21 Mar 2024 14:29:06	U	SMPL
P1804-03 - 1	JC-03-03192024-B	3334	0.4900	-	7471B	PPB	21 Mar 2024 14:31:24	U	SMPL
P1810-01 - 1	342	2020	0.2911	-	7471B	PPB	21 Mar 2024 14:33:42	U	SMPL
CCV99 - 1	CCV99	33427	5.0441	-	7471B	PPB	21 Mar 2024 14:35:59	U	SMPL
CCB99 - 1	CCB99	265	0.0255	-	7471B	PPB	21 Mar 2024 14:38:15	U	SMPL
P1814-01 - 1	TP-6	844	0.1131	-	7471B	PPB	21 Mar 2024 14:40:36	U	SMPL
P1815-01 - 1	OK-01-032024	1759	0.2516	-	7471B	PPB	21 Mar 2024 14:42:51	U	SMPL
P1816-01 - 1	PL-01-032024	1786	0.2557	-	7471B	PPB	21 Mar 2024 14:45:07	U	SMPL
P1817-01 - 1	B-148-SB01	22352	3.3681	-	7471B	PPB	21 Mar 2024 14:47:23	U	SMPL
P1817-02 - 1	B-148-SB03	12405	1.8627	-	7471B	PPB	21 Mar 2024 14:49:38	U	SMPL
P1817-03 - 1	B-148-SB02	2165	0.3130	-	7471B	PPB	21 Mar 2024 14:51:59	U	SMPL
P1786-02LX5 - 1		476	0.0574	-	7471B	PPB	21 Mar 2024 14:54:18	U	SMPL
P1786-02A - 1		29546	4.4568	-	7471B	PPB	21 Mar 2024 14:56:36	U	SMPL
CCV100 - 1	CCV100	33085	4.9924	-	7471B	PPB	21 Mar 2024 15:03:47	U	SMPL
CCB100 - 1	CCB100	546	0.0680	-	7471B	PPB	21 Mar 2024 15:06:02	U	SMPL