



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129981

Ambient Barometric Pressure (mmHg): 750.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	83.0	03/21/2024	09:35
2	P1800-01		1	12	100.00	>212.0	>212.0	03/21/2024	10:00
3	P1800-01DUP		1	12	100.00	>212.0	>212.0	03/21/2024	10:30
4	P1807-05		1	14	100.00	>212.0	>212.0	03/21/2024	11:00
5	P1820-01		1	14	100.00	>212.0	>212.0	03/21/2024	12:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6129981

WorkList Name : FP-0321

WorkList ID : 178761

Department : Wet-Chemistry

Date : 03-21-2024 08:27:29

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1800-01	1394	Water	Flash Point	Cool 4 deg C	PSEG03	O31	03/19/2024	1010B
P1807-05	3462	Water	Flash Point	Cool 4 deg C	PSEG03	O31	03/19/2024	1010B
P1820-01	WASTE-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	P11	03/20/2024	1010B

Date/Time 03/21/2024 09:00
 Raw Sample Received by: RM CWG
 Raw Sample Relinquished by: JPC (007)

Date/Time 03/21/2024 11:10
 Raw Sample Received by: JPC (007)
 Raw Sample Relinquished by: RM CWG