



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129990

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	03/22/2024	08:15
2	CAL2	1	Water	NA	NA	20.3	7.00	03/22/2024	08:16
3	CAL3	1	Water	NA	NA	20.4	10.02	03/22/2024	08:20
4	ICV	1	Water	NA	NA	20.3	7.00	03/22/2024	08:25
5	CCV1	1	Water	NA	NA	20.3	2.01	03/22/2024	08:26
6	P1828-04	1	Water	NA	NA	20.4	6.56	03/22/2024	08:35
7	P1828-04DUP	1	Water	NA	NA	20.5	6.57	03/22/2024	08:36
8	CCV2	1	Water	NA	NA	20.3	12.02	03/22/2024	08:39

WORKLIST(Hardcopy Internal Chain)

W 129990

WorkList Name : ph p1875

WorkList ID : 178775

Department : Wet-Chemistry

Date : 03-22-2024 08:08:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1828-04	TRE-1290	Water	pH	Cool 4 deg C	PSEG03	P21	03/21/2024	9040C

Date/Time 03/22/24 08:10

Raw Sample Received by: JDC

Raw Sample Relinquished by: JDC

Date/Time 03/22/24

Raw Sample Received by: JDC

Raw Sample Relinquished by: JDC

W 129990

JDC

JDC