



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129992

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	03/22/2024	09:55
2	P1799-01		1	12	100.00	>212.0	>212.0	03/22/2024	09:25
3	P1799-01DUP		1	13	100.00	>212.0	>212.0	03/22/2024	09:55
4	P1799-02		1	14	100.00	>212.0	>212.0	03/22/2024	10:25
5	P1799-03		1	12	100.00	>212.0	>212.0	03/22/2024	10:55
6	P1799-04		1	14	100.00	>212.0	>212.0	03/22/2024	11:25
7	P1828-04		1	13	100.00	>212.0	>212.0	03/22/2024	11:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB. 129992

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 03/22/24 flashpoint WorkList ID : 178776 Department : Wet-Chemistry Date : 03-22-2024 08:27:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1799-01	50295	Water	Flash Point	Cool 4 deg C	PSEG03	O31	03/19/2024	1010B
P1799-02	50296	Water	Flash Point	Cool 4 deg C	PSEG03	O31	03/19/2024	1010B
P1799-03	50297	Water	Flash Point	Cool 4 deg C	PSEG03	O31	03/19/2024	1010B
P1799-04	50298	Water	Flash Point	Cool 4 deg C	PSEG03	O31	03/19/2024	1010B
P1828-04	TRE-1290	Water	Flash Point	Cool 4 deg C	PSEG03	P21	03/21/2024	1010B

Date/Time 03/22/2024 9:40
Raw Sample Received by: N.P.W.C
Raw Sample Relinquished by: JP CeeC

Date/Time 03/22/2024 12:25
Raw Sample Received by: JP CeeC
Raw Sample Relinquished by: N.P.W.C