



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129997

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	03/22/2024	12:55
2	P1821-02		1	12	100.00	>212.0	>212.0	03/22/2024	13:25
3	P1821-02DUP		1	13	100.00	>212.0	>212.0	03/22/2024	13:55
4	P1821-03		1	14	100.00	>212.0	>212.0	03/22/2024	14:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-129997

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 03/22/24 flashpointw WorkList ID : 178780 Department : Wet-Chemistry Date : 03-22-2024 11:14:53

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1821-02	HEATER-1-2	Water	Flash Point	Cool 4 deg C	PSEG03	O21	03/20/2024	1010B
P1821-03	HEATER-3	Water	Flash Point	Cool 4 deg C	PSEG03	O21	03/20/2024	1010B

Date/Time 03/22/2024 12:40
Raw Sample Received by: A.P. W.C.
Raw Sample Relinquished by: [Signature]

Date/Time 03/22/2024 14:40
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: N.P. W.C.