

LB130000


SFAM01.1

INSTRUMENT ID: CY1

Linear

 μ Abs.:

47282



A= 0.0000e+000

B= 2.0869e-004 slope

C= -2.0145e-002 y-intercept

Rho= 0.9995046

Accept=Accepted

Max. Conc: —

10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	g/y)
0.00	0.000	-0.003	-0.003	84	0.000	84					-
0.05	0.050					442	0				-
0.20	0.200	0.194	-0.006	1025	0.0 %	1025					-3
2.50	2.500	2.478	-0.022	11969	0.0 %	11969					-1
5.00	5.000	4.950	-0.050	23817	0.0 %	23817					-1
7.50	7.500	7.734	0.234	37154	0.0 %	37154					3
10.0	10.000	9.847	-0.153	47282	0.0 %	47282					-2

LB130000

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	84	-		0 SFAM01.1	PPB	3/22/2024 14:12	Std
	0.2 S01	1025	-		0.2 SFAM01.1	PPB	3/22/2024 14:15	Std
	2.5 S02	11969	-		2.5 SFAM01.1	PPB	3/22/2024 14:17	Std
	5 S03	23817	-		5 SFAM01.1	PPB	3/22/2024 14:19	Std
	7.5 S04	37154	-		7.5 SFAM01.1	PPB	3/22/2024 14:22	Std
	10 S05	47282	-		10 SFAM01.1	PPB	3/22/2024 14:24	Std
ICV060	ICV060	20751	4.3104 -		SFAM01.1	PPB	3/22/2024 14:27	SMPL
ICB060	ICB060	-96	-0.0402 -		SFAM01.1	PPB	3/22/2024 14:29	SMPL
CCV007	CCV007	24224	5.0352 -		SFAM01.1	PPB	3/22/2024 14:31	SMPL
CCB007	CCB007	-79	-0.0366 -		SFAM01.1	PPB	3/22/2024 14:34	SMPL
PB159748BL	PBW748	65	-0.0066 -		SFAM01.1	PPB	3/22/2024 14:36	SMPL
PB159662TB	LEB662	85	-0.0024 -		SFAM01.1	PPB	3/22/2024 14:38	SMPL
P1762-01	E0732	85	-0.0024 -		SFAM01.1	PPB	3/22/2024 14:40	SMPL
P1762-02	E0737	75	-0.0045 -		SFAM01.1	PPB	3/22/2024 14:43	SMPL
P1762-03	E0737D	78	-0.0039 -		SFAM01.1	PPB	3/22/2024 14:45	SMPL
P1762-04	E0737S	4659	0.9522 -		SFAM01.1	PPB	3/22/2024 14:47	SMPL
CCV008	CCV008	23982	4.9847 -		SFAM01.1	PPB	3/22/2024 14:49	SMPL
CCB008	CCB008	-54	-0.0314 -		SFAM01.1	PPB	3/22/2024 14:52	SMPL