



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130007

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	03/22/2024	14:55
2	P1802-01		1	13	100.00	>212.0	>212.0	03/22/2024	15:25
3	P1802-01DUP		1	12	100.00	>212.0	>212.0	03/22/2024	15:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-130007

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 03/22/24 flashpoint o

WorkList ID : 178789

Department : Wet-Chemistry

Date : 03-22-2024 14:04:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1802-01	0116	Solid	Flash Point	Cool 4 deg C	PSEG03	O31	03/19/2024	1010B

Date/Time03/22/202414:45

Raw Sample Received by:N.P.W.C.

Raw Sample Relinquished by:SPCOC

Date/Time03/22/202416:45

Raw Sample Received by:SPCOC

Raw Sample Relinquished by:N.P.W.C.