



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130013

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Ion Chromatography Stock, 500 ml	W2025

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/25/2024	08:00
2	CAL2	1	Water	NA	NA	20.3	7.00	03/25/2024	08:01
3	CAL3	1	Water	NA	NA	20.4	10.02	03/25/2024	08:05
4	ICV	1	Water	NA	NA	20.3	7.02	03/25/2024	08:11
5	CCV1	1	Water	NA	NA	20.4	2.01	03/25/2024	08:15
6	P1837-01	1	Water	NA	NA	20.1	4.55	03/25/2024	08:20
7	P1837-01DUP	1	Water	NA	NA	20.2	4.56	03/25/2024	08:21
8	CCV2	1	Water	NA	NA	20.3	12.02	03/25/2024	08:25

WORKLIST(Hardcopy Internal Chain)

VR130013

WorkList Name : ph p1837

WorkList ID : 178793

Department : Wet-Chemistry

Date : 03-25-2024 07:43:05

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1837-01	GEMINI-POOL-WATER	Water	pH	Cool 4 deg C	STER09	P31	03/15/2024	9040C

Date/Time 03/25/24 07:50
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 03/25/24 10:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]