



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB130014

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Ion Chromatography Stock, 500 ml	W2025

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/25/2024	08:00
2	CAL2	1	Water	NA	NA	20.3	7.00	03/25/2024	08:01
3	CAL3	1	Water	NA	NA	20.4	10.02	03/25/2024	08:05
4	ICV	1	Water	NA	NA	20.3	7.02	03/25/2024	08:11
5	CCV1	1	Water	NA	NA	20.3	2.02	03/25/2024	08:35
6	P1844-03	1	Solid	20.01	20	24.3	4.77	03/25/2024	08:40
7	P1844-03DUP	1	Solid	20.02	20	24.4	4.78	03/25/2024	08:41
8	CCV2	1	Water	NA	NA	20.3	12.02	03/25/2024	08:45

WPS 130014

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity p1844 WorkList ID : 178794 Department : Wet-Chemistry Date : 03-25-2024 07:44:06

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1844-03	BURL-OILY-DEBRIS	Solid	Corrosivity	Cool 4 deg C	PSEG03	P21	03/22/2024	9045D

488
Date/Time 03/25/24 07:55
Raw Sample Received by: 78199
Raw Sample Relinquished by: 517CSM

Date/Time 03/25/24 10:30
Raw Sample Received by: 5DCSM
Raw Sample Relinquished by: 78199