

SFAM01.1


LB130019

INSTRUMENT ID: CV1

Linear

μ Abs.:

44602



A= 0.0000e+000

B= 2.2252e-004 *slope*

C= 9.3811e-003 *intercept*

Rho= 0.9997559

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>0/0.1</i>
0.00	0.000	0.033	0.033	108	0.000	108					=
0.05	0.050					442	0				10
0.20	0.200	0.220	0.020	946	0.0 %	946					-3
2.50	2.500	2.434	-0.066	10895	0.0 %	10895					-2
5.00	5.000	4.924	-0.076	22088	0.0 %	22088					2
7.50	7.500	7.655	0.155	34358	0.0 %	34358					-1
10.0	10.000	9.934	-0.066	44602	0.0 %	44602					

LB130019

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	108	-		0 SFAM01.1	PPB	3/25/2024 13:54	Std
	0.2 S01	946	-		0.2 SFAM01.1	PPB	3/25/2024 13:57	Std
	2.5 S02	10895	-		2.5 SFAM01.1	PPB	3/25/2024 13:59	Std
	5 S03	22088	-		5 SFAM01.1	PPB	3/25/2024 14:01	Std
	7.5 S04	34358	-		7.5 SFAM01.1	PPB	3/25/2024 14:04	Std
	10 S05	44602	-		10 SFAM01.1	PPB	3/25/2024 14:06	Std
ICV063	ICV063	19518	4.3525 -		SFAM01.1	PPB	3/25/2024 14:09	SMPL
ICB063	ICB063	-89	-0.0104 -		SFAM01.1	PPB	3/25/2024 14:11	SMPL
CCV014	CCV014	22578	5.0334 -		SFAM01.1	PPB	3/25/2024 14:13	SMPL
CCB014	CCB014	-74	-0.0071 -		SFAM01.1	PPB	3/25/2024 14:15	SMPL
PB159764BL	PBW764	48	0.0201 -		SFAM01.1	PPB	3/25/2024 14:20	SMPL
P1737-01	YCLE3	847	0.1979 -		SFAM01.1	PPB	3/25/2024 14:23	SMPL
P1737-02	YCLE3D	766	0.1798 -		SFAM01.1	PPB	3/25/2024 14:27	SMPL
P1737-03	YCLE3S	4743	1.0648 -		SFAM01.1	PPB	3/25/2024 14:30	SMPL
P1737-04	YCLE5	183	0.0501 -		SFAM01.1	PPB	3/25/2024 14:32	SMPL
P1737-05	YCLG2	401	0.0986 -		SFAM01.1	PPB	3/25/2024 14:34	SMPL
P1737-06	YCLE7	467	0.1133 -		SFAM01.1	PPB	3/25/2024 14:36	SMPL
P1737-07	YCLE9	638	0.1513 -		SFAM01.1	PPB	3/25/2024 14:39	SMPL
P1737-08	YCLF1	759	0.1783 -		SFAM01.1	PPB	3/25/2024 14:41	SMPL
P1737-09	YCLG4	208	0.0557 -		SFAM01.1	PPB	3/25/2024 14:43	SMPL
P1738-01	YCLE4	571	0.1364 -		SFAM01.1	PPB	3/25/2024 14:46	SMPL
P1738-02	YCLE4D	804	0.1883 -		SFAM01.1	PPB	3/25/2024 14:48	SMPL
P1738-03	YCLE4S	5500	1.2332 -		SFAM01.1	PPB	3/25/2024 14:50	SMPL
P1738-04	YCLE6	342	0.0855 -		SFAM01.1	PPB	3/25/2024 14:52	SMPL
P1738-05	YCLG3	331	0.083 -		SFAM01.1	PPB	3/25/2024 14:55	SMPL
P1738-06	YCLE8	628	0.1491 -		SFAM01.1	PPB	3/25/2024 14:57	SMPL
P1738-07	YCLF0	645	0.1529 -		SFAM01.1	PPB	3/25/2024 14:59	SMPL
P1738-08	YCLF2	841	0.1965 -		SFAM01.1	PPB	3/25/2024 15:01	SMPL
P1738-09	YCLG5	512	0.1233 -		SFAM01.1	PPB	3/25/2024 15:04	SMPL
PB159765BL	PBW765	68	0.0245 -		SFAM01.1	PPB	3/25/2024 15:06	SMPL
CCV015	CCV015	23237	5.18 -		SFAM01.1	PPB	3/25/2024 15:08	SMPL
CCB015	CCB015	-46	-0.0009 -		SFAM01.1	PPB	3/25/2024 15:10	SMPL
P1601-01	MDL-WATER	189	0.0514 -		SFAM01.1	PPB	3/25/2024 15:19	SMPL
CCV016	CCV016	23088	5.1468 -		SFAM01.1	PPB	3/25/2024 15:21	SMPL
CCB016	CCB016	-50	-0.0017 -		SFAM01.1	PPB	3/25/2024 15:24	SMPL