

LB130028

7470A

INSTRUMENT ID: CVL

Linear

 μ Abs.:

41214

A= 0.0000e+000

B= 2.3549e-004

C= 4.1688e-002

Rho= 0.9984081

Accept=Accepted

Max. Conc: —

10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.072	0.072	127	0.000	127					1/100
0.2	0.200	0.179	-0.021	583	0.0 %	583					-11
2.5	2.500	2.406	-0.094	10039	0.0 %	10039					-4
5.0	5.000	4.886	-0.114	20572	0.0 %	20572					-2
7.5	7.500	7.910	0.410	33415	0.0 %	33415					5
10.0	10.000	9.747	-0.253	41214	0.0 %	41214					-3

LB130028 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Mar 2024 09:10:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	127	-	0.0000	7470A	PPB	26 Mar 2024 09:36:05	S	Std
0.2 - 1	SP-2	583	-	0.2000	7470A	PPB	26 Mar 2024 09:38:23	S	Std
2.5 - 1	S2.5	10039	-	2.5000	7470A	PPB	26 Mar 2024 09:40:42	S	Std
5.0 - 1	SS	20572	-	5.0000	7470A	PPB	26 Mar 2024 09:42:59	S	Std
7.5 - 1	S7.5	33415	-	7.5000	7470A	PPB	26 Mar 2024 09:48:04	S	Std
10.0 - 1	S10	41214	-	10.0000	7470A	PPB	26 Mar 2024 09:50:20	S	Std
ICV64 - 1	ICV64	17861	4.2477	-	7470A	PPB	26 Mar 2024 09:59:34	U	SMPL
ICB64 - 1	ICB64	226	0.0949	-	7470A	PPB	26 Mar 2024 10:04:21	U	SMPL
CCV17 - 1	CCV17	20542	4.8791	-	7470A	PPB	26 Mar 2024 10:06:36	U	SMPL
CCB17 - 1	CCB17	55	0.0546	-	7470A	PPB	26 Mar 2024 10:08:51	U	SMPL
CRA - 1	CRA	585	0.1794	-	7470A	PPB	26 Mar 2024 10:11:08	U	SMPL
HighStd - 1	HighStd	40402	9.5558	-	7470A	PPB	26 Mar 2024 10:13:23	U	SMPL
ChkStd - 1	ChkStd	30276	7.1713	-	7470A	PPB	26 Mar 2024 10:15:38	U	SMPL
PB159774BL - 1	PBW	23	0.0471	-	7470A	PPB	26 Mar 2024 10:17:56	U	SMPL
PB159774BS - 1	LCSW	19774	4.6982	-	7470A	PPB	26 Mar 2024 10:20:15	U	SMPL
P1806-01 - 1	389	315	0.1159	-	7470A	PPB	26 Mar 2024 10:22:31	U	SMPL
P1806-01DUP - 1	389DUP	483	0.1554	-	7470A	PPB	26 Mar 2024 10:24:51	U	SMPL
P1806-01MS - 1	389MS	17350	4.1274	-	7470A	PPB	26 Mar 2024 10:29:57	U	SMPL
P1806-01MSD - 1	389MSD	16658	3.9644	-	7470A	PPB	26 Mar 2024 10:32:14	U	SMPL
CCV18 - 1	CCV18	20366	4.8376	-	7470A	PPB	26 Mar 2024 10:34:31	U	SMPL
CCB18 - 1	CCB18	146	0.0761	-	7470A	PPB	26 Mar 2024 10:39:23	U	SMPL
PB159716TB - 1	PB159716TB	118	0.0695	-	7470A	PPB	26 Mar 2024 10:43:30	U	SMPL
P1806-01LX5 - 1		125	0.0711	-	7470A	PPB	26 Mar 2024 10:45:44	U	SMPL
P1806-01A - 1		15758	3.7525	-	7470A	PPB	26 Mar 2024 10:48:00	U	SMPL
CCV19 - 1	CCV19	19481	4.6292	-	7470A	PPB	26 Mar 2024 10:50:15	U	SMPL
CCB19 - 1	CCB19	-44	0.0313	-	7470A	PPB	26 Mar 2024 10:52:33	U	SMPL