

LB130031

7471B INSTRUMENT ID: CY1

Linear

 μ Abs.:

50123

A= 0.0000e+000

B= 2.0025e-004

C= -8.4973e-002

Rho= 0.9999161

Accept=Accepted

Max. Conc:—

10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/ID
0.0	0.000	-0.050	-0.050	174	0.000	174					0/01D
0.2	0.200	0.189	-0.011	1368	0.0 %	1368					6
2.5	2.500	2.525	0.025	13034	0.0 %	13034					1
5.0	5.000	5.090	0.090	25843	0.0 %	25843					2
7.5	7.500	7.494	-0.006	37849	0.0 %	37849					6
10.0	10.000	9.952	-0.048	50123	0.0 %	50123					0

LB130031 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Mar 2024 13:03:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	56	174	-	0.0000	7471B	PPB	26 Mar 2024 13:31:55	S	Std
0.2 - 1	50.2	1368	-	0.2000	7471B	PPB	26 Mar 2024 13:34:12	S	Std
2.5 - 1	12.5	13034	-	2.5000	7471B	PPB	26 Mar 2024 13:36:30	S	Std
5.0 - 1	55	25843	-	5.0000	7471B	PPB	26 Mar 2024 13:38:47	S	Std
7.5 - 1	57.5	37849	-	7.5000	7471B	PPB	26 Mar 2024 13:41:06	S	Std
10.0 - 1	51.0	50123	-	10.0000	7471B	PPB	26 Mar 2024 13:46:07	S	Std
ICV65 - 1	ICV65	19005	3.7207	-	7471B	PPB	26 Mar 2024 13:52:25	U	SMPL
ICB65 - 1	ICB65	119	-0.0611	-	7471B	PPB	26 Mar 2024 13:54:40	U	SMPL
CCV20 - 1	CCV20	25667	5.0548	-	7471B	PPB	26 Mar 2024 13:56:58	U	SMPL
CCB20 - 1	CCB20	60	-0.0730	-	7471B	PPB	26 Mar 2024 13:59:13	U	SMPL
CRA - 1	CRA	1363	0.1880	-	7471B	PPB	26 Mar 2024 14:01:33	U	SMPL
HighStd - 1	HighStd	50454	10.0183	-	7471B	PPB	26 Mar 2024 14:03:48	U	SMPL
ChkStd - 1	ChkStd	35632	7.0502	-	7471B	PPB	26 Mar 2024 14:06:03	U	SMPL
PB159781BS - 1	LCSS	20401	4.0003	-	7471B	PPB	26 Mar 2024 14:10:52	U	SMPL
P1842-01 - 1	EO-01-03222024-A	3002	0.5162	-	7471B	PPB	26 Mar 2024 14:18:36	U	SMPL
P1842-01DUP - 1	EO-01-03222024-ADUP	2753	0.4663	-	7471B	PPB	26 Mar 2024 14:20:52	U	SMPL
P1842-01MS - 1	EO-01-03222024-AMS	24202	4.7614	-	7471B	PPB	26 Mar 2024 14:28:46	U	SMPL
P1842-01MSD - 1	EO-01-03222024-AMSD	23689	4.6587	-	7471B	PPB	26 Mar 2024 14:31:03	U	SMPL
P1842-03 - 1	EO-01-03222024-B	25820	5.0854	-	7471B	PPB	26 Mar 2024 14:33:22	U	SMPL
CCV21 - 1	CCV21	24041	4.7292	-	7471B	PPB	26 Mar 2024 14:35:40	U	SMPL
CCB21 - 1	CCB21	-38	-0.0926	-	7471B	PPB	26 Mar 2024 14:38:02	U	SMPL
P1847-01 - 1	SU-02-032224	3291	0.5740	-	7471B	PPB	26 Mar 2024 14:40:24	U	SMPL
P1861-01 - 1	HD-02-03252024	1169	0.1491	-	7471B	PPB	26 Mar 2024 14:42:39	U	SMPL
P1863-02 - 1	SOIL-PILE-COMP	3074	0.5306	-	7471B	PPB	26 Mar 2024 14:44:57	U	SMPL
P1864-01 - 1	VNJ229-72-11976	2101	0.3357	-	7471B	PPB	26 Mar 2024 14:47:12	U	SMPL
P1865-01 - 1	VNJ-235	28091	5.5401	-	7471B	PPB	26 Mar 2024 14:49:30	U	SMPL
P1865-03 - 1	VNJ-205	753	0.0658	-	7471B	PPB	26 Mar 2024 14:51:49	U	SMPL
PB159781BL - 1	PBS	178	-0.0493	-	7471B	PPB	26 Mar 2024 14:57:30	U	SMPL
P1842-01LX5 - 1		454	0.0059	-	7471B	PPB	26 Mar 2024 14:59:46	U	SMPL
P1842-01A - 1		8586	1.6343	-	7471B	PPB	26 Mar 2024 15:02:03	U	SMPL
P1842-01A - 1		24059	4.7328	-	7471B	PPB	26 Mar 2024 15:12:36	U	SMPL
CCV22 - 1	CCV22	25094	4.9400	-	7471B	PPB	26 Mar 2024 15:17:33	U	SMPL
CCB22 - 1	CCB22	251	-0.0347	-	7471B	PPB	26 Mar 2024 15:19:50	U	SMPL