



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130033

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Ion Chromatography Stock, 500 ml	W2025

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/26/2024	15:05
2	CAL2	1	Water	NA	NA	20.3	7.00	03/26/2024	15:06
3	CAL3	1	Water	NA	NA	20.3	10.02	03/26/2024	15:09
4	ICV	1	Water	NA	NA	20.3	6.99	03/26/2024	15:11
5	CCV1	1	Water	NA	NA	20.3	2.01	03/26/2024	15:15
6	P1878-04	1	Water	NA	NA	20.3	7.37	03/26/2024	15:25
7	P1878-04DUP	1	Water	NA	NA	20.4	7.38	03/26/2024	15:26
8	CCV2	1	Water	NA	NA	20.3	12.02	03/26/2024	15:30

WORKLIST(Hardcopy Internal Chain)

VB 130033

VB

WorkList Name : ph p1878

WorkList ID : 178841

Department : Wet-Chemistry

Date : 03-26-2024 14:55:49

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1878-04	B	E0LZ1		Cool 4 deg C	USEP04	C63	03/20/2024	9040C

Date/Time 03/26/24 15:00

Raw Sample Received by: VB (ccc)

Raw Sample Relinquished by: JT (ccc)

Date/Time 03/26/24

Raw Sample Received by: JT (ccc)

Raw Sample Relinquished by: VB (ccc)

18130

JT (ccc)

VB (ccc)