



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130034

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Ion Chromatography Stock, 500 ml	W2025

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/26/2024	15:05
2	CAL2	1	Water	NA	NA	20.3	7.00	03/26/2024	15:06
3	CAL3	1	Water	NA	NA	20.3	10.02	03/26/2024	15:09
4	ICV	1	Water	NA	NA	20.3	6.99	03/26/2024	15:11
5	CCV1	1	Water	NA	NA	20.4	2.02	03/26/2024	15:40
6	P1878-01	1	Solid	20.02	20	20.6	7.67	03/26/2024	15:45
7	P1878-01DUP	1	Solid	20.03	20	20.7	7.68	03/26/2024	15:46
8	CCV2	1	Water	NA	NA	20.3	12.02	03/26/2024	15:50

WORKLIST(Hardcopy Internal Chain)

130034

WorkList Name : ph p1878s

WorkList ID : 178842

Department : Wet-Chemistry

Date : 03-26-2024 14:56:41

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1878-01	B EOLZ0	Solid	pH	Cool 4 deg C	USEP04	C63	03/20/2024	9045D

Date/Time 03/26/24 15:00
 Raw Sample Received by: JTC
 Raw Sample Relinquished by: JTC

Date/Time 03/26/24 17:30
 Raw Sample Received by: JTC
 Raw Sample Relinquished by: JTC