



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB130042

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Ion Chromatography Stock, 500 ml	W2025

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.00	03/26/2024	13:20
2	CAL2	1	Water	NA	NA	20.4	7.00	03/26/2024	13:21
3	CAL3	1	Water	NA	NA	20.3	10.02	03/26/2024	13:22
4	ICV	1	Water	NA	NA	20.4	7.02	03/26/2024	13:25
5	CCV1	1	Water	NA	NA	20.4	2.01	03/26/2024	13:30
6	P1863-03	1	Solid	20.02	20	22.6	8.11	03/26/2024	13:35
7	P1863-03DUP	1	Solid	20.03	20	22.7	8.12	03/26/2024	13:36
8	P1869-03	1	Solid	20.03	20	23.8	6.43	03/26/2024	13:40
9	P1869-06	1	Solid	20.02	20	23.2	6.30	03/26/2024	13:45
10	P1869-09	1	Solid	20.03	20	23.2	6.20	03/26/2024	13:50
11	P1869-12	1	Solid	20.04	20	22.4	8.15	03/26/2024	14:00
12	P1869-15	1	Solid	20.01	20	21.6	7.83	03/26/2024	14:10
13	P1870-03	1	Solid	20.02	20	21.6	7.64	03/26/2024	14:15
14	CCV2	1	Water	NA	NA	20.3	12.02	03/26/2024	14:25