


LB130049

7471B

INSTRUMENT ID: CVL



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/010
0.0	0.000	-0.089	-0.089	243	0.000	243					(
0.2	0.200	0.198	-0.002	1854	0.0 %	1854					-)
2.5	2.500	2.539	0.039	14992	0.0 %	14992					2
5.0	5.000	5.106	0.106	29399	0.0 %	29399					2
7.5	7.500	7.551	0.051	43119	0.0 %	43119					1
10.0	10.000	9.896	-0.104	56278	0.0 %	56278					-)

LB130049 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Mar 2024 13:15:03

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	243	-	0.000	7471B	PPB	27 Mar 2024 13:33:26	S	Std
0.2 - 1	50.2	1854	-	0.200	7471B	PPB	27 Mar 2024 13:37:57	S	Std
2.5 - 1	52.5	14992	-	2.500	7471B	PPB	27 Mar 2024 13:40:15	S	Std
5.0 - 1	55	29399	-	5.000	7471B	PPB	27 Mar 2024 13:42:32	S	Std
7.5 - 1	57.5	43119	-	7.500	7471B	PPB	27 Mar 2024 13:44:52	S	Std
10.0 - 1	57.0	56278	-	10.000	7471B	PPB	27 Mar 2024 13:52:41	S	Std
ICV66 - 1	ICV66	22669	3.9068	-	7471B	PPB	27 Mar 2024 13:55:50	U	SMPL
ICB66 - 1	ICB66	305	-0.0783	-	7471B	PPB	27 Mar 2024 13:58:05	U	SMPL
CCV23 - 1	CCV23	28335	4.9165	-	7471B	PPB	27 Mar 2024 14:00:25	U	SMPL
CCB23 - 1	CCB23	308	-0.0777	-	7471B	PPB	27 Mar 2024 14:02:40	U	SMPL
CRA - 1	CRA	1731	0.1759	-	7471B	PPB	27 Mar 2024 14:05:36	U	SMPL
HighStd - 1	HighStd	54108	9.5090	-	7471B	PPB	27 Mar 2024 14:07:51	U	SMPL
ChkStd - 1	ChkStd	37505	6.5505	-	7471B	PPB	27 Mar 2024 14:10:06	U	SMPL
PB159788BL - 1	PB159788BL	252	-0.0877	-	7471B	PPB	27 Mar 2024 14:12:34	U	SMPL
PB159788BS - 1	PB159788BS	20305	3.4856	-	7471B	PPB	27 Mar 2024 14:15:55	U	SMPL
P1187-01 - 1	LOD-MDL-SOIL-03-QT1-2024	1059	0.0561	-	7471B	PPB	27 Mar 2024 14:40:46	U	SMPL
P1187-01RE - 1	LOD-MDL-SOIL-03-QT1-2024RE	1460	0.1276	-	7471B	PPB	27 Mar 2024 14:43:02	U	SMPL
P1187-03 - 1	MDL-SOIL-03-QT1-2024	1047	0.0540	-	7471B	PPB	27 Mar 2024 14:53:05	U	SMPL
P1869-01 - 1	SOIL-1	4917	0.7436	-	7471B	PPB	27 Mar 2024 14:55:22	U	SMPL
P1869-01DUP - 1	SOIL-1DUP	4814	0.7252	-	7471B	PPB	27 Mar 2024 14:57:37	U	SMPL
CCV24 - 1	CCV24	30660	5.3308	-	7471B	PPB	27 Mar 2024 14:59:54	U	SMPL
CCB24 - 1	CCB24	429	-0.0562	-	7471B	PPB	27 Mar 2024 15:02:11	U	SMPL
P1869-01MS - 1	SOIL-1MS	28120	4.8782	-	7471B	PPB	27 Mar 2024 15:04:33	U	SMPL
P1869-01MSD - 1	SOIL-1MSD	29294	5.0874	-	7471B	PPB	27 Mar 2024 15:06:48	U	SMPL
P1869-04 - 1	SOIL-2	4546	0.6775	-	7471B	PPB	27 Mar 2024 15:09:07	U	SMPL
P1869-07 - 1	SOIL-3	3051	0.4111	-	7471B	PPB	27 Mar 2024 15:11:27	U	SMPL
P1869-10 - 1	SOIL-4	3492	0.4896	-	7471B	PPB	27 Mar 2024 15:13:44	U	SMPL
P1869-13 - 1	SOIL-5	2756	0.3585	-	7471B	PPB	27 Mar 2024 15:16:01	U	SMPL
P1870-01 - 1	EXTRA-PILE	4292	0.6322	-	7471B	PPB	27 Mar 2024 15:18:19	U	SMPL
P1869-01LX5 - 1		704	-0.0072	-	7471B	PPB	27 Mar 2024 15:20:36	U	SMPL
P1869-01A - 1		27719	4.8067	-	7471B	PPB	27 Mar 2024 15:22:54	U	SMPL
CCV25 - 1	CCV25	31475	5.4760	-	7471B	PPB	27 Mar 2024 15:25:10	U	SMPL
CCB25 - 1	CCB25	431	-0.0558	-	7471B	PPB	27 Mar 2024 15:27:29	U	SMPL