



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130063

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Ion Chromatography Stock, 500 ml	W2025

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/27/2024	16:25
2	CAL2	1	Water	NA	NA	20.3	6.99	03/27/2024	16:26
3	CAL3	1	Water	NA	NA	20.3	10.02	03/27/2024	16:30
4	ICV	1	Water	NA	NA	20.4	7.02	03/27/2024	16:33
5	CCV1	1	Water	NA	NA	20.3	2.01	03/27/2024	16:35
6	P1856-04	1	Water	NA	NA	20.3	7.37	03/27/2024	16:40
7	P1875-01	1	Water	NA	NA	20.3	3.92	03/27/2024	16:49
8	P1901-04	1	Water	NA	NA	20.3	7.28	03/27/2024	16:55
9	P1901-04DUP	1	Water	NA	NA	20.4	7.29	03/27/2024	16:56
10	CCV2	1	Water	NA	NA	20.3	12.02	03/27/2024	16:59

130063

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH P1901 WorkList ID : 178900 Department : Wet-Chemistry Date : 03-27-2024 16:17:12

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1856-04	P E0LZ1	Water	pH	Cool 4 deg C	USEP04	A11	03/20/2024	9040C
P1875-01	A MEG-DRUM	Water	pH	Cool 4 deg C	PSEG03	K11	03/26/2024	9040C
P1901-04	A A069	Water	pH	Cool 4 deg C	PSEG03	P21	03/27/2024	9040C

Date/Time 03/27/24 16:20
Raw Sample Received by: 26 C00C1
Raw Sample Relinquished by: [Signature]

Date/Time 03/27/24 17:40
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: 26 C00C1