

L8130066

SFAM01.1

INSTRUMENT ID: CY1

Linear

μ Abs.:

51483

A= 0.0000e+000

B= 1.9448e-004 *slope*

C= -6.0270e-002 *intercept*

Rho= 0.9998527

Accept=Accepted

Max. Conc:—

10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>Q/A</i>
0.00	0.000	0.045	0.045	540	0.000	540					<i>2 3</i>
0.05	0.050					442	0				<i>5</i>
0.20	0.200	0.205	0.005	1365	0.0 %	1365					<i>1</i>
2.50	2.500	2.386	-0.114	12578	0.0 %	12578					<i>1</i>
5.00	5.000	5.041	0.041	26228	0.0 %	26228					<i>0</i>
7.50	7.500	7.571	0.071	39241	0.0 %	39241					
10.0	10.000	9.952	-0.048	51483	0.0 %	51483					

LB130066

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	540	-		0 SFAM01.1	PPB	3/28/2024 10:56	Std
	0.2 S01	1365	-		0.2 SFAM01.1	PPB	3/28/2024 10:58	Std
	2.5 S02	12578	-		2.5 SFAM01.1	PPB	3/28/2024 11:00	Std
	5 S03	26228	-		5 SFAM01.1	PPB	3/28/2024 11:03	Std
	7.5 S04	39241	-		7.5 SFAM01.1	PPB	3/28/2024 11:06	Std
	10 S05	51483	-		10 SFAM01.1	PPB	3/28/2024 11:11	Std
ICV067	ICV067	21804	4.1802 -		SFAM01.1	PPB	3/28/2024 11:13	SMPL
ICB067	ICB067	540	0.0447 -		SFAM01.1	PPB	3/28/2024 11:17	SMPL
CCV026	CCV026	26380	5.0701 -		SFAM01.1	PPB	3/28/2024 11:19	SMPL
CCB026	CCB026	430	0.0234 -		SFAM01.1	PPB	3/28/2024 11:22	SMPL
PB159826BL	PBS826	251	-0.0115 -		SFAM01.1	PPB	3/28/2024 11:27	SMPL
P1889-01	E0794	1075	0.1488 -		SFAM01.1	PPB	3/28/2024 11:29	SMPL
P1889-02	E0795	1067	0.1472 -		SFAM01.1	PPB	3/28/2024 11:31	SMPL
P1889-03	E0796	1249	0.1826 -		SFAM01.1	PPB	3/28/2024 11:33	SMPL
P1889-04	E0797	1882	0.3057 -		SFAM01.1	PPB	3/28/2024 11:36	SMPL
P1889-05	E0798	1918	0.3127 -		SFAM01.1	PPB	3/28/2024 11:38	SMPL
P1889-06	E07A0	1717	0.2737 -		SFAM01.1	PPB	3/28/2024 11:40	SMPL
P1889-07	E07A0D	1625	0.2558 -		SFAM01.1	PPB	3/28/2024 11:42	SMPL
P1889-08	E07A0S	14194	2.7002 -		SFAM01.1	PPB	3/28/2024 11:45	SMPL
P1890-01	E0799	655	0.0671 -		SFAM01.1	PPB	3/28/2024 11:47	SMPL
P1890-02	E0799D	649	0.0659 -		SFAM01.1	PPB	3/28/2024 11:49	SMPL
P1890-03	E0799S	14732	2.8048 -		SFAM01.1	PPB	3/28/2024 11:52	SMPL
PB159827BL	PBS827	114	-0.0381 -		SFAM01.1	PPB	3/28/2024 11:54	SMPL
P1601-03	MDL-SOIL	623	0.0609 -		SFAM01.1	PPB	3/28/2024 12:02	SMPL
CCV027	CCV027	25441	4.8875 -		SFAM01.1	PPB	3/28/2024 12:04	SMPL
CCB027	CCB027	416	0.0206 -		SFAM01.1	PPB	3/28/2024 12:07	SMPL