



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130068

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	03/28/2024	10:50
2	P1875-01		1	8	20.00	68.0	67.0	03/28/2024	11:20
3	P1901-04		1	14	100.00	>212.0	>212.0	03/28/2024	11:50
4	P1901-04DUP		1	13	100.00	>212.0	>212.0	03/28/2024	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-130068

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 03/28/24 flashpoint WorkList ID : 178921 Department : Wet-Chemistry Date : 03-28-2024 10:34:35

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1875-01	MEG-DRUM	Water	Flash Point	Cool 4 deg C	PSEG03	K11	03/26/2024	1010B
P1901-04	A069	Water	Flash Point	Cool 4 deg C	PSEG03	P21	03/27/2024	1010B

Date/Time 03/28/2024 10:45
Raw Sample Received by: N.P.W.C
Raw Sample Relinquished by: Jd GOC

Date/Time 03/28/2024 12:50
Raw Sample Received by: Jd GOC
Raw Sample Relinquished by: N.P.W.C