

LB130076

7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

64157

A= 0.0000e+000

B= 1.5668e-004

C= 5.1891e-002

Rho= 0.9997646

Accept=Accepted

Max. Conc:—

10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9/11
0.0	0.000	0.099	0.099	302	0.000	302					0
0.2	0.200	0.204	0.004	969	0.0 %	969	0				2
2.5	2.500	2.464	-0.036	15393	0.0 %	15393					-1
5.0	5.000	4.884	-0.116	30838	0.0 %	30838					-2
7.5	7.500	7.446	-0.054	47193	0.0 %	47193					-1
10.0	10.000	10.104	0.104	64157	0.0 %	64157					1

LB130076 INSTRUMENT ID ;; CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Mar 2024 10:19:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	302	-	0.0000	7470A	PPB	29 Mar 2024 10:41:28	S	Std
0.2 - 1	50.2	969	-	0.2000	7470A	PPB	29 Mar 2024 10:45:52	S	Std
2.5 - 1	52.5	15393	-	2.5000	7470A	PPB	29 Mar 2024 10:48:09	S	Std
5.0 - 1	55	30838	-	5.0000	7470A	PPB	29 Mar 2024 10:50:26	S	Std
7.5 - 1	57.5	47193	-	7.5000	7470A	PPB	29 Mar 2024 10:52:43	S	Std
10.0 - 1	510	64157	-	10.0000	7470A	PPB	29 Mar 2024 11:01:14	S	Std
ICV68 - 1	ICV68	24361	3.8687	-	7470A	PPB	29 Mar 2024 11:09:34	U	SMPL
ICB68 - 1	ICB68	290	0.0973	-	7470A	PPB	29 Mar 2024 11:14:00	U	SMPL
CCV28 - 1	CCV28	28847	4.5716	-	7470A	PPB	29 Mar 2024 11:23:50	U	SMPL
CCB28 - 1	CCB28	21	0.0552	-	7470A	PPB	29 Mar 2024 11:26:05	U	SMPL
CRA - 1	CRA	772	0.1728	-	7470A	PPB	29 Mar 2024 11:28:23	U	SMPL
HighStd - 1	HighStd	66613	10.4887	-	7470A	PPB	29 Mar 2024 11:34:09	U	SMPL
ChkStd - 1	ChkStd	41137	6.4972	-	7470A	PPB	29 Mar 2024 11:39:01	U	SMPL
PB159845BL - 1	PBW	4	0.0525	-	7470A	PPB	29 Mar 2024 11:41:17	U	SMPL
PB159845BS - 1	LCSW	21537	3.4263	-	7470A	PPB	29 Mar 2024 11:43:35	U	SMPL
P1830-07 - 1	VNJ-206	183	0.0806	-	7470A	PPB	29 Mar 2024 11:45:51	U	SMPL
P1830-08 - 1	VNJ-226	310	0.1005	-	7470A	PPB	29 Mar 2024 11:48:10	U	SMPL
P1869-03 - 1	SOIL-1	333	0.1041	-	7470A	PPB	29 Mar 2024 11:50:27	U	SMPL
P1869-06 - 1	SOIL-2	292	0.0976	-	7470A	PPB	29 Mar 2024 11:54:04	U	SMPL
P1869-09 - 1	SOIL-3	356	0.1077	-	7470A	PPB	29 Mar 2024 11:56:18	U	SMPL
CCV29 - 1	CCV29	32988	5.2204	-	7470A	PPB	29 Mar 2024 11:58:35	U	SMPL
CCB29 - 1	CCB29	147	0.0749	-	7470A	PPB	29 Mar 2024 12:03:23	U	SMPL
P1869-12 - 1	SOIL-4	382	0.1117	-	7470A	PPB	29 Mar 2024 12:05:38	U	SMPL
P1869-15 - 1	SOIL-5	350	0.1067	-	7470A	PPB	29 Mar 2024 12:07:53	U	SMPL
P1870-03 - 1	EXTRA-PILE	228	0.0876	-	7470A	PPB	29 Mar 2024 12:10:08	U	SMPL
P1870-03DUP - 1	EXTRA-PILEDUP	243	0.0900	-	7470A	PPB	29 Mar 2024 12:12:24	U	SMPL
P1870-03MS - 1	EXTRA-PILEMS	17372	2.7737	-	7470A	PPB	29 Mar 2024 12:14:39	U	SMPL
P1870-03MSD - 1	EXTRA-PILEMSD	18476	2.9467	-	7470A	PPB	29 Mar 2024 12:16:56	U	SMPL
PB159846BL - 1	PBW	87	0.0655	-	7470A	PPB	29 Mar 2024 12:19:47	U	SMPL
PB159846BS - 1	LCSW	23402	3.7185	-	7470A	PPB	29 Mar 2024 12:25:43	U	SMPL
P1883-02 - 1	DOMINICK-STREET	239	0.0893	-	7470A	PPB	29 Mar 2024 12:28:00	U	SMPL
P1894-01 - 1	COMP-1	348	0.1064	-	7470A	PPB	29 Mar 2024 12:30:17	U	SMPL
CCV30 - 1	CCV30	31521	4.9905	-	7470A	PPB	29 Mar 2024 12:32:33	U	SMPL
CCB30 - 1	CCB30	-16	0.0494	-	7470A	PPB	29 Mar 2024 12:34:49	U	SMPL
P1894-02 - 1	COMP-2	356	0.1077	-	7470A	PPB	29 Mar 2024 12:37:07	U	SMPL
P1894-03 - 1	COMP-3	344	0.1058	-	7470A	PPB	29 Mar 2024 12:39:22	U	SMPL
P1894-03DUP - 1	COMP-3DUP	369	0.1097	-	7470A	PPB	29 Mar 2024 12:41:38	U	SMPL
P1894-03MS - 1	COMP-3MS	26528	4.2083	-	7470A	PPB	29 Mar 2024 12:46:34	U	SMPL
P1894-03MSD - 1	COMP-3MSD	25647	4.0702	-	7470A	PPB	29 Mar 2024 12:48:50	U	SMPL
PB159868BL - 1	PBW	59	0.0611	-	7470A	PPB	29 Mar 2024 12:51:09	U	SMPL
PB159868BS - 1	LCSW	22022	3.5023	-	7470A	PPB	29 Mar 2024 12:58:25	U	SMPL
P1187-09 - 1	MDL-WATER-03-QT1-2024	114	0.0698	-	7470A	PPB	29 Mar 2024 13:05:48	U	SMPL
P1867-04 - 1	B-148-GW01	405	0.1153	-	7470A	PPB	29 Mar 2024 13:08:04	U	SMPL
P1867-04DUP - 1	B-148-GW01DUP	410	0.1161	-	7470A	PPB	29 Mar 2024 13:10:20	U	SMPL
CCV31 - 1	CCV31	28678	4.5451	-	7470A	PPB	29 Mar 2024 13:21:40	U	SMPL
CCB31 - 1	CCB31	-58	0.0428	-	7470A	PPB	29 Mar 2024 13:23:55	U	SMPL
P1867-04MS - 1	B-148-GW01MS	18681	2.9788	-	7470A	PPB	29 Mar 2024 13:26:13	U	SMPL
P1867-04MSD - 1	B-148-GW01MSD	18772	2.9931	-	7470A	PPB	29 Mar 2024 13:28:29	U	SMPL
P1867-05 - 1	EB-03252024	163	0.0774	-	7470A	PPB	29 Mar 2024 13:30:46	U	SMPL
P1901-04 - 1	A069	419	0.1175	-	7470A	PPB	29 Mar 2024 13:33:04	U	SMPL
P1187-07 - 1	MDL-WATER	46	0.0591	-	7470A	PPB	29 Mar 2024 13:35:20	U	SMPL
P1187-07RE - 1	MDL-WATERRE	521	0.1335	-	7470A	PPB	29 Mar 2024 13:40:13	U	SMPL
PB159778TB - 1	PB159778TB	79	0.0643	-	7470A	PPB	29 Mar 2024 13:47:27	U	SMPL
PB159823TB - 1	PB159823TB	81	0.0646	-	7470A	PPB	29 Mar 2024 13:49:44	U	SMPL
CCV32 - 1	CCV32	29299	4.6424	-	7470A	PPB	29 Mar 2024 13:54:33	U	SMPL
CCB32 - 1	CCB32	-64	0.0419	-	7470A	PPB	29 Mar 2024 13:56:49	U	SMPL
P1870-03LX5 - 1		100	0.0676	-	7470A	PPB	29 Mar 2024 13:59:07	U	SMPL
P1870-03A - 1		18571	2.9616	-	7470A	PPB	29 Mar 2024 14:01:22	U	SMPL
P1894-03LX5 - 1		-12	0.0500	-	7470A	PPB	29 Mar 2024 14:03:38	U	SMPL
P1894-03A - 1		20308	3.2337	-	7470A	PPB	29 Mar 2024 14:05:56	U	SMPL
P1867-04LX5 - 1		-16	0.0494	-	7470A	PPB	29 Mar 2024 14:08:11	U	SMPL
P1867-04A - 1		18645	2.9732	-	7470A	PPB	29 Mar 2024 14:10:29	U	SMPL
CCV33 - 1	CCV33	28900	4.5799	-	7470A	PPB	29 Mar 2024 14:12:47	U	SMPL
CCB33 - 1	CCB33	-61	0.0423	-	7470A	PPB	29 Mar 2024 14:15:05	U	SMPL