



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130081

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Ion Chromatography Stock, 500 ml	W2025

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/29/2024	08:05
2	CAL2	1	Water	NA	NA	20.3	7.00	03/29/2024	08:06
3	CAL3	1	Water	NA	NA	20.3	10.02	03/29/2024	08:10
4	ICV	1	Water	NA	NA	20.3	7.01	03/29/2024	08:15
5	CCV1	1	Water	NA	NA	20.6	2.02	03/29/2024	08:55
6	P1856-01	1	Solid	20.02	20	20.6	7.67	03/29/2024	09:00
7	P1915-01	1	Solid	20.03	20	22.4	9.83	03/29/2024	09:10
8	P1915-01DUP	1	Solid	20.04	20	22.5	9.84	03/29/2024	09:11
9	CCV2	1	Water	NA	NA	20.3	12.02	03/29/2024	09:15

WORKLIST(Hardcopy Internal Chain)

UP130081

WorkList Name : ph p1916

WorkList ID : 178935

Department : Wet-Chemistry

Date : 03-29-2024 07:39:55

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1856-01	E0LZ0	Solid	pH	Cool 4 deg C	USEP04	A11	03/20/2024	9045D
P1915-01	WC-A-COMP	Solid	pH	Cool 4 deg C	EART12	I22	03/28/2024	9045D

Date/Time 03/29/24 07:45
 Raw Sample Received by: JH (Q2C)
 Raw Sample Relinquished by: JDCSM

Date/Time 03/29/24
 Raw Sample Received by: JDCSM
 Raw Sample Relinquished by: JH (Q2C)