



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB130094

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP107199
Turbidity Calibration std, 20NTU	WP107194
Turbidity Calibration std, 40NTU	WP107193
Turbidity Calibration std, 1NTU	WP107197
Turbidity Calibration std, 5NTU	WP107196
Turbidity Calibration - CCV std, 10 NTU	WP107195
10 NTU Standard 500 ml	W2971

Intercept: 0.1801

Slope: 1.0437

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.156	-0.02		03/29/2024	09:30
2	CAL2	1	1	1.260	1.04	3.5	03/29/2024	09:33
3	CAL3	5	1	5.290	4.90	-2.1	03/29/2024	09:36
4	CAL4	10	1	11.100	10.46	4.6	03/29/2024	09:40
5	CAL5	20	1	20.500	19.47	-2.7	03/29/2024	09:43
6	CAL6	40	1	42.100	40.17	0.4	03/29/2024	09:46



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB130094

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.500	9.888	03/29/2024	09:50
2	ICB		1	0.168	-0.012	03/29/2024	09:53
3	CCV1	10	1	10.800	10.175	03/29/2024	09:56
4	CCB1		1	0.168	-0.012	03/29/2024	09:53
5	LB130094BL		1	0.170	-0.010	03/29/2024	10:03
6	P1911-34		1	15.400	14.583	03/29/2024	10:06
7	CCV2	10	1	10.900	10.271	03/29/2024	10:26
8	CCB2		1	0.172	-0.008	03/29/2024	10:30