



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB130095

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP107199
Turbidity Calibration std, 20NTU	WP107194
Turbidity Calibration std, 40NTU	WP107193
Turbidity Calibration std, 1NTU	WP107197
Turbidity Calibration std, 5NTU	WP107196
Turbidity Calibration - CCV std, 10 NTU	WP107195
10 NTU Standard 500 ml	W2971

Intercept: 0.1801

Slope: 1.0437

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.156	-0.02		03/29/2024	09:30
2	CAL2	1	1	1.260	1.04	3.5	03/29/2024	09:33
3	CAL3	5	1	5.290	4.90	-2.1	03/29/2024	09:36
4	CAL4	10	1	11.100	10.46	4.6	03/29/2024	09:40
5	CAL5	20	1	20.500	19.47	-2.7	03/29/2024	09:43
6	CAL6	40	1	42.100	40.17	0.4	03/29/2024	09:46



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Iwona

Parameter: Turbidity

SUPERVISOR REVIEW BY: Sohil

Run Number: LB130095

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.500	9.888	03/29/2024	09:50
2	ICB		1	0.168	-0.012	03/29/2024	09:53
3	CCV1	10	1	10.800	10.175	03/29/2024	09:56
4	CCB1		1	0.164	-0.015	03/29/2024	10:00
5	LB130095BL		1	0.170	-0.010	03/29/2024	10:03
6	P1911-33		1	15.400	14.583	03/29/2024	10:06
7	P1916-01		1	2.050	1.792	03/29/2024	10:10
8	P1916-01DUP		1	2.020	1.763	03/29/2024	10:13
9	P1916-03		1	0.712	0.510	03/29/2024	10:16
10	P1916-05		1	0.536	0.341	03/29/2024	10:20
11	P1916-07		1	0.705	0.503	03/29/2024	10:23
12	CCV2	10	1	10.900	10.271	03/29/2024	10:26
13	CCB2		1	0.172	-0.008	03/29/2024	10:30

WORKLIST(Hardcopy Internal Chain)

LB 130095

WorkList Name : TURB-032824

WorkList ID : 178930

Department : Wet-Chemistry

Date : 03-28-2024 14:49:16

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1911-33	PT-TURB-WP	Water	Turbidity	Cool 4 deg C	CHEM02	QA Of	03/26/2024	SM2130 B
P1911-34	PT-TURB-WP	Water	Turbidity	Cool 4 deg C	CHEM02	QA Of	03/26/2024	180.1
P1916-01	MW-1	Water	Turbidity	Cool 4 deg C	LOCK01	I23	03/27/2024	SM2130 B
P1916-03	MW-3	Water	Turbidity	Cool 4 deg C	LOCK01	I23	03/27/2024	SM2130 B
P1916-05	MW-4	Water	Turbidity	Cool 4 deg C	LOCK01	I23	03/27/2024	SM2130 B
P1916-07	MW-2	Water	Turbidity	Cool 4 deg C	LOCK01	I23	03/27/2024	SM2130 B

Date/Time 03/29/24 09:10
 Raw Sample Received by: 12 (JC)
 Raw Sample Relinquished by: 78 (CO)

Date/Time 03/29/24 10:45
 Raw Sample Received by: 78 (CO)
 Raw Sample Relinquished by: 12 (JC)