


LB130096

7471B

INSTRUMENT ID: 081

Linear



A= 0.0000e+000
B= 1.8979e-004
C= -1.0931e-002
Rho= 0.9999396
Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.024	0.024	186	0.000	186					
0.2	0.200	0.183	-0.017	1024	0.0 %	1024					-9
2.5	2.500	2.532	0.032	13398	0.0 %	13398					'
5.0	5.000	4.970	-0.030	26245	0.0 %	26245					-1
7.5	7.500	7.436	-0.064	39237	0.0 %	39237					-1
10.0	10.000	10.054	0.054	53035	0.0 %	53035					'

LB130096 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 29 Mar 2024 15:55:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	186	-	0.0000	7471B	PPB	31 Mar 2024 13:26:36	S	Std
0.2 - 1	50-2	1024	-	0.2000	7471B	PPB	31 Mar 2024 13:29:28	S	Std
2.5 - 1	52.5	13398	-	2.5000	7471B	PPB	31 Mar 2024 13:31:46	S	Std
5.0 - 1	55	26245	-	5.0000	7471B	PPB	31 Mar 2024 13:34:02	S	Std
7.5 - 1	57.5	39237	-	7.5000	7471B	PPB	31 Mar 2024 13:39:05	S	Std
10.0 - 1	510	53035	-	10.0000	7471B	PPB	31 Mar 2024 13:44:10	S	Std
ICV69 - 1	ICV69	20457	3.8716	-	7471B	PPB	31 Mar 2024 13:47:03	U	SMPL
ICB69 - 1	ICB69	-42	-0.0189	-	7471B	PPB	31 Mar 2024 13:49:19	U	SMPL
CCV34 - 1	CCV34	26043	4.9317	-	7471B	PPB	31 Mar 2024 13:51:38	U	SMPL
CCB34 - 1	CCB34	-82	-0.0265	-	7471B	PPB	31 Mar 2024 13:53:53	U	SMPL
CRA - 1	CRA	1111	0.1999	-	7471B	PPB	31 Mar 2024 14:00:22	U	SMPL
HighStd - 1	HighStd	53033	10.0541	-	7471B	PPB	31 Mar 2024 14:07:04	U	SMPL
ChkStd - 1	ChkStd	41389	7.8442	-	7471B	PPB	31 Mar 2024 14:09:21	U	SMPL
PB159877BL - 1	PBS	62	0.0008	-	7471B	PPB	31 Mar 2024 14:11:47	U	SMPL
PB159877BS - 1	LCSS	20715	3.9205	-	7471B	PPB	31 Mar 2024 14:14:11	U	SMPL
P1874-01 - 1	VNJ-209	30488	5.7753	-	7471B	PPB	31 Mar 2024 14:16:27	U	SMPL
P1874-03 - 1	72-11933	24650	4.6674	-	7471B	PPB	31 Mar 2024 14:18:47	U	SMPL
P1876-01 - 1	TR-06-032624	2134	0.3941	-	7471B	PPB	31 Mar 2024 14:21:09	U	SMPL
P1877-01 - 1	TRE-1291	1343	0.2440	-	7471B	PPB	31 Mar 2024 14:23:30	U	SMPL
P1880-01 - 1	YANUZZI-GROUP	273947	51.9810	-	7471B	PPB	31 Mar 2024 14:25:46	U	SMPL
CCV35 - 1	CCV35	25954	4.9148	-	7471B	PPB	31 Mar 2024 14:33:29	U	SMPL
CCB35 - 1	CCB35	247	0.0359	-	7471B	PPB	31 Mar 2024 14:36:25	U	SMPL
P1883-01 - 1	DOMINICK-STREET	331544	62.9123	-	7471B	PPB	31 Mar 2024 14:38:41	U	SMPL
P1894-04 - 1	RB40858RT	4772	0.8947	-	7471B	PPB	31 Mar 2024 14:40:57	U	SMPL
P1894-06 - 1	RT3414	1632	0.2988	-	7471B	PPB	31 Mar 2024 14:43:43	U	SMPL
P1894-06DUP - 1	RT3414DUP	1947	0.3586	-	7471B	PPB	31 Mar 2024 14:46:02	U	SMPL
P1894-06MS - 1	RT3414MS	21911	4.1475	-	7471B	PPB	31 Mar 2024 14:48:18	U	SMPL
P1894-06MSD - 1	RT3414MSD	24556	4.6495	-	7471B	PPB	31 Mar 2024 14:50:35	U	SMPL
P1900-01 - 1	SS-1	3621	0.6763	-	7471B	PPB	31 Mar 2024 14:52:56	U	SMPL
P1913-01 - 1	CCM-1	452	0.0749	-	7471B	PPB	31 Mar 2024 14:55:16	U	SMPL
P1917-01 - 1	CCM-2	991	0.1771	-	7471B	PPB	31 Mar 2024 14:57:34	U	SMPL
CCV36 - 1	CCV36	27998	5.3028	-	7471B	PPB	31 Mar 2024 14:59:50	U	SMPL
CCB36 - 1	CCB36	155	0.0185	-	7471B	PPB	31 Mar 2024 15:02:06	U	SMPL
P1894-06KLX5 - 1		375	0.0602	-	7471B	PPB	31 Mar 2024 15:04:25	U	SMPL
P1894-06A - 1		22477	4.2549	-	7471B	PPB	31 Mar 2024 15:06:41	U	SMPL
P1880-01DLX10 - 1		29503	5.5884	-	7471B	PPB	31 Mar 2024 15:12:27	U	SMPL
P1883-01DLX10 - 1		26739	5.0638	-	7471B	PPB	31 Mar 2024 15:14:42	U	SMPL
CCV37 - 1	CCV37	24979	4.7298	-	7471B	PPB	31 Mar 2024 15:17:03	U	SMPL
CCB37 - 1	CCB37	-1	-0.0111	-	7471B	PPB	31 Mar 2024 15:19:23	U	SMPL