



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB130097

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Ion Chromatography Stock, 500 ml	W2025

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	03/29/2024	15:47
2	CAL2	1	Water	NA	NA	20.3	6.99	03/29/2024	15:50
3	CAL3	1	Water	NA	NA	20.3	10.02	03/29/2024	15:55
4	ICV	1	Water	NA	NA	20.3	7.01	03/29/2024	16:00
5	CCV1	1	Water	NA	NA	20.2	2.02	03/29/2024	16:05
6	P1935-03	1	Water	NA	NA	22.8	12.02	03/29/2024	16:15
7	P1937-01	1	Water	NA	NA	21.1	6.65	03/29/2024	16:25
8	P1937-01DUP	1	Water	NA	NA	21.2	6.66	03/29/2024	16:26
9	CCV2	1	Water	NA	NA	20.3	12.02	03/29/2024	16:35

WORKLIST(Hardcopy Internal Chain)

VB 130097

WorkList Name : PH P1975

WorkList ID : 178972

Department : Wet-Chemistry

Date : 03-29-2024 15:31:13

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1935-03	BELL-TOTES	Water	pH	Cool 4 deg C	PSEG03	I23	03/29/2024	9040C
P1937-01	A5552	Water	pH	Cool 4 deg C	PSEG03	P31	03/29/2024	9040C

Date/Time 03/29/24 15:40

Raw Sample Received by: JG

Raw Sample Relinquished by: JG

Date/Time 03/29/24

Raw Sample Received by: JG

Raw Sample Relinquished by: JG

14:00

JG

JG