

SFAM01.1


LB130101

INSTRUMENT ID: CV2

Linear

 μ Abs.:

49554



A= 0.0000e+000

B= 2.0692e-004 slope

C= -3.8924e-002 intercept

Rho= 0.9991644

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.00	0.000	0.076	0.076	554	0.000	554					
0.05	0.050					442	0				11
0.20	0.200	0.209	0.009	1199	0.0 %	1199					5
2.50	2.500	2.587	0.087	12689	0.0 %	12689					3
5.00	5.000	4.842	-0.158	23589	0.0 %	23589					3
7.50	7.500	7.272	-0.228	35331	0.0 %	35331					3
10.0	10.000	10.215	0.215	49554	0.0 %	49554					2

LB130101

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	554	-		0 SFAM01.1	PPB	3/31/2024 18:59	Std
	0.2 S01	1199	-		0.2 SFAM01.1	PPB	3/31/2024 19:02	Std
	2.5 S02	12689	-		2.5 SFAM01.1	PPB	3/31/2024 19:04	Std
	5 S03	23589	-		5 SFAM01.1	PPB	3/31/2024 19:14	Std
	7.5 S04	35331	-		7.5 SFAM01.1	PPB	3/31/2024 19:16	Std
	10 S05	49554	-		10 SFAM01.1	PPB	3/31/2024 19:21	Std
ICV071	ICV071	20923	4.2904 -		SFAM01.1	PPB	3/31/2024 19:24	SMPL
ICB071	ICB071	412	0.0463 -		SFAM01.1	PPB	3/31/2024 19:26	SMPL
CCV040	CCV040	23460	4.8154 -		SFAM01.1	PPB	3/31/2024 19:29	SMPL
CCB040	CCB040	371	0.0378 -		SFAM01.1	PPB	3/31/2024 19:31	SMPL
PB159901BL	PBW901	142	-0.0095 -		SFAM01.1	PPB	3/31/2024 19:36	SMPL
P1932-09	E07D2	430	0.0501 -		SFAM01.1	PPB	3/31/2024 19:38	SMPL
PB159902BL	PBW902	83	-0.0218 -		SFAM01.1	PPB	3/31/2024 19:41	SMPL
P1601-02	MDL-WATER	417	0.0474 -		SFAM01.1	PPB	3/31/2024 19:46	SMPL
CCV041	CCV041	23207	4.763 -		SFAM01.1	PPB	3/31/2024 19:48	SMPL
CCB041	CCB041	424	0.0488 -		SFAM01.1	PPB	3/31/2024 19:50	SMPL