


LB130105

7471B

INSTRUMENT ID: CV1

 μ Abs.:

69209



Linear

A= 0.0000e+000

B= 1.4253e-004

C= 2.0503e-002

Rho= 0.9995386

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	U/D
0.0	0.000	0.057	0.057	258	0.000	258					
0.2	0.200	0.196	-0.004	1229	0.0 %	1229					2
2.5	2.500	2.409	-0.091	16761	0.0 %	16761					4
5.0	5.000	4.937	-0.063	34494	0.0 %	34494					1
7.5	7.500	7.716	0.216	53992	0.0 %	53992					3
10.0	10.000	9.885	-0.115	69209	0.0 %	69209					1

LB130105 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 01 Apr 2024 11:54:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SP	258	-	0.0000	7471B	PPB	01 Apr 2024 12:41:44	S	Std
0.2 - 1	SP-2	1229	-	0.2000	7471B	PPB	01 Apr 2024 12:44:01	S	Std
2.5 - 1	SP-5	16761	-	2.5000	7471B	PPB	01 Apr 2024 12:46:19	S	Std
5.0 - 1	SP	34494	-	5.0000	7471B	PPB	01 Apr 2024 12:48:36	S	Std
7.5 - 1	SP-5	53992	-	7.5000	7471B	PPB	01 Apr 2024 12:56:13	S	Std
10.0 - 1	SP-10	69209	-	10.0000	7471B	PPB	01 Apr 2024 13:08:38	S	Std
ICV72 - 1	ICV72	26358	3.7773	-	7471B	PPB	01 Apr 2024 13:13:31	U	SMPL
ICB72 - 1	ICB72	292	0.0621	-	7471B	PPB	01 Apr 2024 13:20:09	U	SMPL
CCV42 - 1	CCV42	34399	4.9234	-	7471B	PPB	01 Apr 2024 13:22:24	U	SMPL
CCB42 - 1	CCB42	0	0.0205	-	7471B	PPB	01 Apr 2024 13:24:43	U	SMPL
CRA - 1	CRA	1221	0.1945	-	7471B	PPB	01 Apr 2024 13:27:03	U	SMPL
HighStd - 1	HighStd	69538	9.9317	-	7471B	PPB	01 Apr 2024 13:29:20	U	SMPL
ChkStd - 1	ChkStd	53131	7.5932	-	7471B	PPB	01 Apr 2024 13:31:39	U	SMPL
PB159876BL - 1	PBS	-194	-0.0071	-	7471B	PPB	01 Apr 2024 13:34:06	U	SMPL
PB159876BS - 1	LCSS	23167	3.3225	-	7471B	PPB	01 Apr 2024 13:36:32	U	SMPL
P1879-01 - 1	EB-01(0-2)	3102	0.4626	-	7471B	PPB	01 Apr 2024 13:43:55	U	SMPL
P1879-01DUP - 1	EB-01(0-2)DUP	2905	0.4345	-	7471B	PPB	01 Apr 2024 13:46:11	U	SMPL
P1879-01MS - 1	EB-01(0-2)MS	27713	3.9704	-	7471B	PPB	01 Apr 2024 13:48:29	U	SMPL
P1879-01MSD - 1	EB-01(0-2)MSD	26901	3.8547	-	7471B	PPB	01 Apr 2024 13:50:47	U	SMPL
P1879-02 - 1	EB-01(2-4)	26828	3.8443	-	7471B	PPB	01 Apr 2024 13:53:56	U	SMPL
CCV43 - 1	CCV43	33980	4.8636	-	7471B	PPB	01 Apr 2024 13:56:49	U	SMPL
CCB43 - 1	CCB43	18	0.0231	-	7471B	PPB	01 Apr 2024 14:01:18	U	SMPL
P1879-03 - 1	EB-02(0-2)	4588	0.6744	-	7471B	PPB	01 Apr 2024 14:03:33	U	SMPL
P1879-04 - 1	EB-02(2-4)	6011	0.8772	-	7471B	PPB	01 Apr 2024 14:05:48	U	SMPL
P1879-05 - 1	EB-03(0-2)	7583	1.1013	-	7471B	PPB	01 Apr 2024 14:08:05	U	SMPL
P1879-06 - 1	EB-03(2-4)	6012	0.8774	-	7471B	PPB	01 Apr 2024 14:10:23	U	SMPL
P1879-07 - 1	EB-04(0-2)	3408	0.5062	-	7471B	PPB	01 Apr 2024 14:12:41	U	SMPL
P1879-08 - 1	EB-04(2-4)	4764	0.6995	-	7471B	PPB	01 Apr 2024 14:14:59	U	SMPL
P1879-09 - 1	HS-01(0-6)	3531	0.5238	-	7471B	PPB	01 Apr 2024 14:17:17	U	SMPL
P1879-10 - 1	HS-02(0-6)	2768	0.4150	-	7471B	PPB	01 Apr 2024 14:19:37	U	SMPL
P1879-11 - 1	HS-03(0-6)	6585	0.9591	-	7471B	PPB	01 Apr 2024 14:21:54	U	SMPL
P1879-12 - 1	HS-04(0-6)	9900	1.4315	-	7471B	PPB	01 Apr 2024 14:24:11	U	SMPL
CCV44 - 1	CCV44	36557	5.2309	-	7471B	PPB	01 Apr 2024 14:30:01	U	SMPL
CCB44 - 1	CCB44	3	0.0209	-	7471B	PPB	01 Apr 2024 14:35:06	U	SMPL
P1915-01 - 1	WC-A-COMP	18046	2.5926	-	7471B	PPB	01 Apr 2024 14:37:22	U	SMPL
P1927-01 - 1	72-11991	1116	0.1796	-	7471B	PPB	01 Apr 2024 14:39:37	U	SMPL
P1927-03 - 1	72-11998	1074	0.1736	-	7471B	PPB	01 Apr 2024 14:41:56	U	SMPL
P1936-01 - 1	TR-04-032924	2097	0.3194	-	7471B	PPB	01 Apr 2024 14:47:35	U	SMPL
P1937-03 - 1	369	953	0.1563	-	7471B	PPB	01 Apr 2024 14:49:51	U	SMPL
P1938-01 - 1	PAT-02-032924	4395	0.6469	-	7471B	PPB	01 Apr 2024 14:53:03	U	SMPL
P1879-01LX5 - 1		430	0.0818	-	7471B	PPB	01 Apr 2024 14:55:19	U	SMPL
P1879-01A - 1		27868	3.9925	-	7471B	PPB	01 Apr 2024 14:57:38	U	SMPL
CCV45 - 1	CCV45	31624	4.5278	-	7471B	PPB	01 Apr 2024 15:08:10	U	SMPL
CCB45 - 1	CCB45	-233	-0.0127	-	7471B	PPB	01 Apr 2024 15:10:25	U	SMPL