



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB130112

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3080

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	04/01/2024	13:15
2	P1935-03		1	12	100.00	>212.0	>212.0	04/01/2024	13:45
3	P1935-03DUP		1	13	100.00	>212.0	>212.0	04/01/2024	14:15
4	P1937-01		1	14	100.00	>212.0	>212.0	04/01/2024	14:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-130112

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 04/01/24 flashpoint WorkList ID : 179026 Department : Wet-Chemistry Date : 04-01-2024 12:33:36

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1935-03	BELL-TOTES	Water	Flash Point	Cool 4 deg C	PSEG03	I23	03/29/2024	1010B
P1937-01	A5552	Water	Flash Point	Cool 4 deg C	PSEG03	P31	03/29/2024	1010B

Date/Time 04/01/2024 13:00
Raw Sample Received by: N.P.W.C.
Raw Sample Relinquished by: J. G. C.

Date/Time 04/01/2024 15:40
Raw Sample Received by: J. G. C.
Raw Sample Relinquished by: N.P.W.C.